

Deciding When to Submit a 510(k) for a Change to an Existing Device

Draft Guidance for Industry and Food and Drug Administration Staff

DRAFT GUIDANCE

This draft guidance document is being distributed for comment purposes only.

Document issued on August 8, 2016.

You should submit comments and suggestions regarding this draft document within 90 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <http://www.regulations.gov>. Submit written comments to the Division of Dockets Management (HFA-305), Food and Drug Administration, 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document regarding CDRH-regulated devices, contact the 510(k) Staff at 301-796-5640.

For questions regarding the use or interpretation of this guidance in the review of submissions to the Center for Biologics Evaluation and Research, contact the Office of Communication, Outreach and Development at 1-800-835-4709 or 240-402-8010 or by email at ocod@fda.hhs.gov.

When final, this document will supersede *Deciding When to Submit a 510(k) for a Change to an Existing Device*, dated January 10, 1997.



**U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health**

Center for Biologics Evaluation and Research

Contains Nonbinding Recommendations

Draft – Not for Implementation

Preface

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an e-mail request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please use the document number 1500054 to identify the guidance you are requesting.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), by written request, Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov or from the Internet at <http://www.fda.gov/BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

Table of Contents

| | |
|--|----|
| 1. Introduction | 1 |
| 2. Background | 1 |
| 3. Scope | 3 |
| 4. Guiding Principles | 4 |
| 5. How to Use This Guidance | 6 |
| A. Labeling Changes | 11 |
| B. Technology, Engineering, and Performance Changes | 18 |
| C. Materials Changes | 27 |
| D. Technology, Engineering, Performance, and Materials Changes for In Vitro Diagnostic Devices | 31 |
| E. Considerations for Risk Assessments of Modified Devices | 36 |
| Appendix A: Examples | 39 |
| Appendix B: Documentation | 58 |
| Appendix C: Definitions | 69 |

Flowcharts

| | |
|---------------------------|----|
| Figure 1 - Main Flowchart | 8 |
| Figure 2 - Flowchart A | 12 |
| Figure 3 - Flowchart B | 19 |
| Figure 4 - Flowchart C | 28 |
| Figure 5 - Flowchart D | 32 |

Deciding When to Submit a 510(k) for a Change to an Existing Device

Draft Guidance for Industry and Food and Drug Administration Staff

This guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff responsible for this guidance as listed on the title page.

1. Introduction

Almost from the enactment of the Medical Device Amendments to the Federal Food, Drug, and Cosmetic Act (the FD&C Act) in 1976, the Food and Drug Administration (FDA or the Agency) has attempted to define with greater clarity when a change in a medical device would trigger the requirement that a [manufacturer](#) submit a new premarket notification (510(k)) to the Agency. When finalized, this document will supersede [Deciding When to Submit a 510\(k\) for a Change to an Existing Device \(K97-1\)](#), issued on January 10, 1997.

For the current edition of the FDA-recognized standards referenced in this document, see the FDA Recognized Consensus Standards Database at <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>.

FDA's guidance documents, including this draft guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency's current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidance means that something is suggested or recommended, but not required.

2. Background

Contains Nonbinding Recommendations

Draft – Not for Implementation

36 The regulatory criteria in 21 CFR 807.81(a)(3) state that a premarket notification must be
37 submitted when:

38

39 *(3) The device is one that the person currently has in commercial distribution or is*
40 *reintroducing into commercial distribution, but that is about to be significantly*
41 *changed or modified in design, components, method of manufacture, or intended use.*
42 *The following constitute significant changes or modifications that require a*
43 *premarket notification:*

44

45 *(i) A change or modification in the device that could significantly affect the*
46 *safety or effectiveness of the device, e.g., a significant change or modification*
47 *in design, material, chemical composition, energy source, or manufacturing*
48 *process.*

49

50 *(ii) A major change or modification in the intended use of the device.*

51

52 FDA issued the original [Deciding When to Submit a 510\(k\) for a Change to an Existing](#)
53 [Device \(K97-1\)](#) on January 10, 1997 to provide guidance on this regulatory language. As
54 stated in that guidance, the key issue regarding 21 CFR 807.81(a)(3) is that the phrase “could
55 significantly affect the safety or effectiveness of the device” and the use of the adjectives
56 "major" and "significant" sometimes lead FDA and device manufacturers to different
57 interpretations. That guidance provided the Agency’s interpretation of these terms, with
58 principles and points for manufacturers to consider in analyzing how changes in devices may
59 affect safety or effectiveness and determining whether a new 510(k) must be submitted for a
60 particular type of change. This draft guidance preserves the basic format and content of the
61 original, with updates to add clarity. The added clarity is intended to increase consistent
62 interpretations of the guidance by FDA staff and manufacturers.

63

64 **The 510(k) Process and the Quality System Regulation**

65

66 Any guidance on 510(k)s for changes to a legally marketed device should consider the role
67 the Quality System (QS) regulation, 21 CFR Part 820, plays in changes to devices. For some
68 types of changes to a device, the Agency believes that a new 510(k) is not necessary and that
69 reliance on existing QS requirements may reasonably assure the safety and effectiveness of
70 the changed device.

71

72 Among other requirements, the QS regulation requires manufacturers of finished medical
73 devices to review and approve changes to device design and production (21 CFR 820.30 and
74 820.70) and document changes and approvals in the device master record (21 CFR 820.181).
75 Any process whose results cannot be fully verified by subsequent inspection and testing must
76 be validated (21 CFR 820.75), and changes to the process require review, evaluation, and
77 revalidation of the process where appropriate (21 CFR 820.75(c)).

78

79 The net effect of the QS regulation is to require that, when manufacturers of a finished
80 medical device make a change in the design of a device, there is a process in place to

Contains Nonbinding Recommendations

Draft – Not for Implementation

81 demonstrate that the manufactured device meets the change in design specifications (or the
82 original specifications, if no change was intended). They must keep records, and these
83 records must be made available to an FDA investigator (see Section 704(e) of the FD&C
84 Act). For many types of changes to a device, a new 510(k) may not be required per 21 CFR
85 807.81(a)(3). In these cases, compliance with the QS regulation can reasonably assure the
86 safety and effectiveness of the changed device.

87 **3. Scope**

88
89 This guidance, when finalized, will aid manufacturers of medical devices subject to
90 premarket notification requirements who intend to modify a 510(k)-cleared device or a
91 [preamendments device](#) subject to 510(k) (also referred to together as “existing devices”)
92 during the process of deciding whether the modification exceeds the regulatory threshold of
93 21 CFR 807.81(a)(3) for submission and clearance of a new 510(k). Note that any person
94 required to register under 21 CFR 807.20 who plans to introduce a device into commercial
95 distribution for the first time must, per 21 CFR 807.81(a)(2), submit a 510(k) if that device is
96 not exempt from premarket notification requirements.¹ This guidance, when finalized, is not
97 intended to address modifications to devices that are 510(k)-exempt or require premarket
98 approval (PMA).

99
100 This document incorporates concepts and recommendations from existing FDA guidance and
101 policy, such as [Submission and Review of Sterility Information in Premarket Notification](#)
102 [\(510\(k\)\) Submissions for Devices Labeled as Sterile](#), and device-specific guidance documents
103 regarding when new 510(k)s are required based on modifications to an existing device. In
104 some cases, FDA’s thinking has derived from its experience in situations involving only a
105 few manufacturers of a limited number of devices. In such instances, we have attempted to
106 generalize the concepts to apply to a broader range of devices. However, special cases exist
107 where FDA has established definitive guidance for modifications to specific devices, e.g.,
108 FDA’s guidance on daily wear contact lenses, [Premarket Notification \(510\(k\)\) Guidance](#)
109 [Document for Daily Wear Contact Lenses](#). This guidance, when finalized, is not intended to
110 supersede such device-specific guidance but may cover areas not addressed in such device-
111 specific guidance.

112
113 Recalls: This guidance, when finalized, is also intended to apply to situations when a legally
114 marketed existing device is the subject of a recall and a change in the device or its [labeling](#) is
115 indicated. For more information on recommended procedures in a recall situation, please see
116 Blue Book Memorandum K95-1, [510\(k\) Requirements During Firm-Initiated Recalls](#). As
117 stated in that guidance, if a correction alters a device rather than simply restoring it to its
118 original specifications, a new 510(k) may be required. This guidance, when finalized, may be

¹ Also note that devices with changes requiring a new 510(k) may not be legally commercially distributed before FDA clears the changed device. See 21 CFR 801.100(a) and sections 513(f)(1) and 513(i) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

119 useful in determining whether a new 510(k) is warranted in cases where the correction does
120 alter the device.

121

122 Private Label Distributors and Repackagers: Private label distributors and repackagers are
123 exempt from submitting a 510(k) if they satisfy the requirements of 21 CFR 807.85(b).

124

125 Software Changes: This draft guidance does not apply to [software](#) changes or modifications,
126 however, this guidance does apply to non-software changes to devices containing software or
127 software that is a medical device on its own. Labeling changes to existing devices that
128 contain or consist of software are covered by Section A of this guidance, and non-software
129 technology changes and materials changes to existing devices that contain software are
130 covered by Sections B through D of this guidance. FDA is issuing [a separate draft guidance](#)
131 [document on software](#) changes or modifications concurrently with this draft guidance.

132

133 Combination Products: This draft guidance does not specifically address combination
134 products, such as drug/device or biologic/device combinations, however, the general
135 principles and concepts described herein may be helpful to manufacturers in determining
136 whether a 510(k) is necessary for changes to device constituent parts of combination
137 products. Furthermore, this guidance, when finalized, is not intended to address whether
138 510(k) submissions are required from remanufacturers of existing devices who do not [hold](#)
139 [the 510\(k\)](#) for the device, such as reprocessors of [single-use devices](#). Remanufacturer is
140 defined at 21 CFR 820.3(w) as “any person who processes, conditions, renovates,
141 repackages, restores, or does any other act to a finished device that significantly changes the
142 finished device’s performance or safety specifications, or [intended use](#).”

143

144 4. Guiding Principles

145

146 In using this guidance for deciding whether to submit a new 510(k) for a modification to an
147 existing device, a number of guiding principles should be followed. Some derive from
148 existing FDA 510(k) policy and are widely known, and others are necessary for using the
149 logic scheme contained in this guidance. Thus, anyone using this guidance should bear in
150 mind the following guiding principles:

151

- 152 • **Modifications made with intent to significantly affect safety or effectiveness of a**
153 **device** – If a manufacturer modifies their device with the intent to significantly
154 improve the safety or effectiveness of the device (for example, in response to a
155 known [risk](#), adverse events, etc.), a new 510(k) is likely required. Changes that are
156 not intended to significantly affect the safety or effectiveness of a device, however,
157 should still be evaluated to determine whether the change could significantly affect
158 device safety or effectiveness.
- 159 ○ If a manufacturer modifies their device to address a violation or recall, they
160 should refer to FDA guidances Blue Book Memorandum K95-1, [510\(k\)](#)
161

Contains Nonbinding Recommendations

Draft – Not for Implementation

[Requirements During Firm-Initiated Recalls](#) and [Distinguishing Medical Device Recalls from Medical Device Enhancements](#).

- 162
163
164
165 • **“Could significantly affect” evaluation and the role of testing** – To determine
166 whether a change or modification could significantly affect the safety or effectiveness
167 of a device, the manufacturer should first conduct a risk-based assessment, using the
168 guidance below, of whether the change could significantly affect the device’s safety
169 or effectiveness, either positively or negatively. This risk-based assessment should
170 identify and analyze all new risks and changes in known risks resulting from the
171 device modification, and lead to an initial decision whether or not a new 510(k) is
172 required. If the initial decision following the risk assessment is that a new 510(k) is
173 not required, this decision should be confirmed by successful, routine verification and
174 validation activities. If routine verification and validation activities produce any
175 unexpected issues, any prior decision that a new 510(k) is not required should be
176 reconsidered, as discussed in **B5.4** for non-IVD devices and **D4** for IVD devices.
177 Verification and validation requirements apply for all devices subject to 21 CFR
178 820.30.
179
- 180 • **Unintended consequences of changes** – In evaluating whether a change requires a
181 new 510(k), manufacturers should consider whether there are any unintended
182 consequences or effects of the device change. For example, changes in [sterilization](#)
183 may unintentionally affect device materials, or changes to materials may
184 unintentionally affect the performance of the device.
185
- 186 • **Use of risk management** – The risk profile referred to throughout this document is
187 based on the combination of multiple risk concepts which are important for managing
188 the risks of medical devices. [Hazards](#) and hazardous situations, risk estimation, risk
189 acceptability, risk control, risk/benefit analysis and overall risk evaluation are all
190 concepts that can be applied during the design and development of a medical device.
191 The concept of risk, as defined in ISO 14971: *Medical devices – Application of risk*
192 *management to medical devices*, is the combination of the probability of occurrence
193 of [harm](#) and the severity of that harm. Although the risk terminology used in this
194 document is primarily derived from ISO 14971, it is recognized that an individual
195 manufacturer’s terminology may differ. Because 21 CFR 807.81(a)(3)(i) requires a
196 new 510(k) when a change “could significantly affect safety or effectiveness,” both
197 safety and effectiveness should be considered in evaluating a device’s risk profile, as
198 explained in Section E.
199
- 200 • **Evaluating simultaneous changes** – Because many simultaneous changes may be
201 considered at once, each change should be assessed separately, as well as in
202 aggregate.
203
- 204 • **Appropriate comparative device and cumulative effect of changes** – In using this
205 guidance to help determine whether a particular change requires the submission of a
206 new 510(k), manufacturers should make a risk-based assessment that compares the

Contains Nonbinding Recommendations

Draft – Not for Implementation

207 changed device to their device as previously found to be substantially equivalent in
208 their most recently cleared 510(k) (or to their preamendments device, if no 510(k) has
209 been cleared). Manufacturers may make a number of changes without having to
210 submit a new 510(k), but each time they make a change, the device they should
211 compare it to is their most recently cleared device. When the manufacturer compares
212 the proposed modified device to the device in its most recently cleared 510(k), the
213 manufacturer should evaluate the cumulative impact of all changes since their most
214 recently cleared 510(k).

215

216 • **Documentation requirement** – Whenever manufacturers change their device, they
217 must take certain actions to comply with the QS regulation, 21 CFR Part 820, unless
218 the device in question is exempt by regulation from the QS regulation. The QS
219 regulation requires, among other things, that device changes be documented (See
220 Appendix B for recommendations on [documentation](#)).

221

222 • **510(k) submissions for modified devices** – When a new 510(k) is submitted for a
223 device with multiple modifications, that 510(k) should describe all changes that
224 trigger the requirement for a new 510(k). That 510(k) should also describe other
225 modifications since the last cleared 510(k) (i.e., those that did not require a new
226 510(k)) that would have been documented as part of the original 510(k) for that
227 device. This helps ensure that FDA has a more complete understanding of the device
228 under review. For instance, an original 510(k) would not typically identify or describe
229 individual components of a circuit board, such as resistors, and therefore FDA would
230 not expect modifications to the resistors to be listed in the new 510(k) for a modified
231 device if they did not trigger the requirement for a 510(k). However, 510(k)s typically
232 include a listing of device [warnings](#) in the labeling, so if a warning in the device’s
233 labeling has been modified, that change should be described in the new 510(k) even if
234 that change did not itself trigger the requirement for a new 510(k).

235

236 ○ If a manufacturer makes multiple changes to a device, but only one change
237 triggers the requirement for a new 510(k), the changes that do not require a
238 new 510(k) may be immediately implemented, so long as those changes can
239 be implemented independently of changes that do require a new 510(k). Those
240 changes should, however, be described in the new 510(k) for the change that
241 does require submission (subject to the preceding bullet).

242

243 • **Substantial equivalence determinations** – Manufacturers should understand that,
244 even though they may follow this guidance and submit a new 510(k), a substantially
245 equivalent determination is not assured. See [The 510\(k\) Program: Evaluating
246 Substantial Equivalence in Premarket Notifications \(510\(k\)\)](#) for more information on
247 the decision-making process FDA uses to determine substantial equivalence.

248

249 5. How to Use This Guidance

Contains Nonbinding Recommendations

Draft – Not for Implementation

250

251 This guidance uses flowcharts and text to guide manufacturers through the logic scheme we
252 recommend to arrive at a decision on whether to submit a new 510(k) for a change to an
253 existing device. A single logic scheme containing all the necessary steps would be large and
254 cumbersome and could be quite daunting. Therefore, one is not included in this document.
255 Rather, for ease of use, the single scheme has been broken down into smaller sections that
256 include:

257

- 258 • The main types of changes that might be made to a device (this section, Main
259 Flowchart)
- 260 • Labeling changes (Section A, Flowchart A)
- 261 • Technology, engineering, and performance changes (Section B, Flowchart B)
- 262 • Materials changes (Section C, Flowchart C)
- 263 • Technology, engineering, performance, and materials changes for [in vitro diagnostic](#)
264 [devices](#) (IVDs) (Section D, Flowchart D)
- 265 • Considerations for risk assessments of modified devices (Section E)

266

267 Note that sections B and C are only applicable to non-IVDs, and section D is only applicable
268 to IVDs. All other sections apply to IVDs and non-IVDs alike.


269

270 Please refer to Appendix C: Definitions, for the meaning of terms used in the guidance,
271 including in the flowcharts.

272

Contains Nonbinding Recommendations

Draft – Not for Implementation



273
274

Figure 1 - Main Flowchart

Contains Nonbinding Recommendations

Draft – Not for Implementation

275

276 **Manufacturers should use the flowcharts in concert with the guiding principles above,**
277 **the recommendations in the sections below, and the examples provided in Appendix A.**

278 Answer the questions posed in the text for each individual type of change (e.g., performance
279 change, material change) until a decision is made either to submit a new 510(k) or to
280 document the basis for concluding that a new 510(k) is not required. As mentioned above,
281 when making the decision on whether to submit a new 510(k) for changes, the manufacturer's
282 basis for comparison of any changed device should be the device described in the
283 manufacturer's most recently cleared 510(k) for this device, or to their legally marketed
284 preamendments device. Manufacturers are required to submit a new 510(k) when a change
285 (or changes) exceeds the §807.81(a)(3) threshold, "could significantly affect the safety or
286 effectiveness of the device," or constitutes a "major change or modification in the intended
287 use of the device." This significant effect could be positive or negative. One must keep in
288 mind that what may on the surface appear to be one discrete change to a device may involve
289 multiple changes of various types.

290

291 **Although this guidance does not specifically discuss manufacturing changes, a**
292 **manufacturer should consider the impact of all manufacturing changes on device**
293 **labeling, technology/performance, and/or materials.** If the manufacturing change affects
294 any of these three areas, manufacturers should evaluate the impact of the resulting labeling,
295 technology/performance, or material change using the appropriate flowcharts and companion
296 text. **In cases with multiple changes, manufacturers should use all applicable flowcharts**
297 **and explanatory text.** Consider the following examples:

298

299 Example 1: Multiple changes caused by a manufacturing process change

300

301 A manufacturer decides to change the manufacturing process for a patient-contacting
302 part from a machining process to a stamping process. The use of the stamping process
303 requires a change in the grade of stainless steel and also results in a change of the
304 dimensional tolerances. To evaluate the impact of this change, the manufacturer
305 should use both Sections B (Technology, Engineering, and Performance) and C
306 (Materials).

307

308 Example 2: Multiple changes related to a change in [shelf-life](#)

309

310 A manufacturer changes one or more materials in a device to improve the shelf-life of
311 the product. The material change also affects some of the performance characteristics,
312 resulting in the need to update the labeling. To evaluate the impact of the change, the
313 manufacturer should use Sections A (Labeling), B (Technology, Engineering, and
314 Performance) and C (Materials) or D (Technology, Engineering, Performance, and
315 Materials Changes for IVD Devices).

316

317 For those circumstances where the proposed change is not addressed in this guidance or in a
318 device-specific guidance document, manufacturers are encouraged to contact [CDRH staff or](#)
319 [CBER staff](#).

Contains Nonbinding Recommendations

Draft – Not for Implementation

320

321 Note that the flowchart entries, “new 510(k)” and “documentation,” are written in this way
322 only for conciseness. The reader should interpret “new 510(k)” as **a new 510(k) is likely**
323 **required** and “documentation” as **a new 510(k) is likely not required, document your**
324 **analysis and file it for future reference.**

325

326 Each of the questions listed on the detailed flowcharts are identified by the flowchart letter
327 (A through D) and a sequential number. Those questions on the main spine of the flowcharts
328 relate to major questions to be answered. Subsidiary questions are identified by the flowchart
329 letter, the question number, a decimal point, and another sequential number (e.g., B4.1 is a
330 decision point containing a follow-up question that builds off a determination made in
331 decision point B4).

332

333 Note that the first question is always whether the change is being made with the intent to
334 significantly improve the safety or effectiveness of the device, for example, in response to a
335 known risk, adverse event, etc. (Figure 1 – Main Flowchart). If so, then the change likely
336 “could significantly affect safety or effectiveness” and a new 510(k) likely must be
337 submitted.

338

339 This guidance provides a logic scheme that incorporates risk assessment for evaluating
340 specific types of device changes and modifications, and, in instances where it is not possible
341 to provide further specific guidance, refers to Section E, which provides recommendations
342 for how manufacturers should utilize risk management principles to evaluate their own
343 specific changes and modifications. Because 21 CFR 807.81(a)(3)(i) requires a new 510(k)
344 when a change “could significantly affect safety or effectiveness,” both safety and
345 effectiveness should be considered in evaluating a device’s risk profile, as explained in
346 Section E.

347

348

349 **A. Labeling Changes**

350

351 As noted above, the types of changes are divided into labeling changes, technology,
352 engineering, or performance changes, and materials changes. All labeling changes should be
353 evaluated using a separate logic scheme that concentrates on changes in [indications for use](#)
354 for determining whether clearance of a new 510(k) is required. Other labeling changes are
355 more frequently recommended for documentation only.

356

357 Flowchart A describes the logic scheme to be used when determining when a new 510(k) is
358 required for a labeling change. Changes in device labeling often pose the most difficult
359 questions to be addressed by device manufacturers when deciding when a new 510(k) is
360 required. Frequently, an apparently subtle change in a device’s labeling can have a
361 significant impact on the safe and effective use of the device.

362

363 Confusion often results when discussing the distinction between “indications for use” and the
364 “intended use” of the device. For purposes of substantial equivalence, the term intended use
365 means the general purpose of the device or its function, and encompasses the indications for
366 use.² The indications for use describe the disease or condition the device will diagnose, treat,
367 prevent, cure or mitigate, including a description of the patient population for which the
368 device is intended.³ The indications include all the labeled patient uses of the device. The
369 concept of intended use has particular relevance in determining whether a device can be
370 cleared for marketing through the premarket notification (510(k)) process or must be
371 evaluated in a premarket approval application (PMA) or a *de novo* request for classification
372 under section 513(f)(2) of the FD&C Act. Manufacturers should recognize that, per section
373 513(i) of the FD&C Act, if a particular labeling change results in a different intended use of
374 the device, the device would not be substantially equivalent and a PMA or a *de novo*
375 submission would be required to market the device.

376

377 Rather than referring to “intended use” as a determinant in deciding when to submit a new
378 510(k), this guidance identifies several types of labeling changes or modifications that have a
379 major impact on intended use and thus would require the submission of a new 510(k).⁴ FDA
380 interprets major changes in intended use to be a type of change that could significantly affect
381 safety or effectiveness.

² See FDA’s guidance, [The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \(510\(k\)\)](#).

³ *Ibid.*

⁴ Labeling changes are not the only type of changes that could result in a major change in intended use. See 21 CFR 801.4.

Contains Nonbinding Recommendations

Draft – Not for Implementation




Figure 2 - Flowchart A

382
383
384
385
386
387
388
389
390
391
392
393
394
395
396
397

A1. Is it a substantive change in the indications for use? Changes in the indications for use section of labeling raise more Agency concern than any other aspect of labeling. In fact, most changes in this part of the labeling that affect the substance, meaning, or scope of the indications for use – referred to here as “substantive changes” – could significantly affect safety or effectiveness and will require the submission of a new 510(k). Changes that clarify the indications without affecting the substance or meaning of the indications usually do not require a new 510(k). In addition, some changes in the indications for use that limit use within the currently cleared indication may occur without the submission of a new 510(k). For example, if a device was cleared for use with three specific indications and the firm decides to market the device for only two of those indications, this change would not likely require submission of a new 510(k).

Contains Nonbinding Recommendations

Draft – Not for Implementation

398 Common changes to the indications for use that typically could significantly affect
399 safety or effectiveness and therefore usually require submission of a new 510(k) are:

400

- 401 (1) [Reuse](#) of devices previously labeled “single-use only.”
- 402 (2) Changes from prescription to over the counter (OTC).
- 403 (3) Changes that introduce a new therapeutic or diagnostic claim.
- 404 (4) Changes to allow device use in a new patient population.
- 405 (5) Changes to the type of joint, organ, bone, vasculature, or tissue applied
406 to or interacted with.

407

408 Common changes that likely would not constitute a major change in intended use and
409 would not require a new 510(k) include:

410

- 411 (1) Changes to the device name or description that are consistent with the
412 cleared indications for use; and
- 413 (2) Changes to improve readability or clarity that do not affect the
414 substance of the indications for use.

415 Whether other indication changes require a new 510(k) will be more dependent on the
416 specific device, the original indications for use, and the modified indications for use.
417 To determine whether such types of changes to the indications for use could
418 significantly affect the device’s safety or effectiveness, manufacturers should
419 consider how the change affects the device’s risk profile. As discussed in Section E, a
420 change that introduces a new risk or significantly modifies an existing risk likely
421 requires a new 510(k). The following are examples of types of indication for use
422 changes that may require a new 510(k), as well as points to consider in determining
423 whether a new 510(k) is required:

424

- 425 (1) Changes in use environment.
 - 426 • How a change of this type affects a device’s risk profile depends
427 on the differences in use environment and [environmental](#)
428 [specifications](#). For example, a change from use in a surgical suite
429 to use in a hospital recovery room, both of which will have
430 professional healthcare supervision, may not affect the device’s
431 risk profile. Changes from professional use to home use⁵ or
432 hospital use to ambulatory transport, however, are more likely to
433 affect the device’s risk profile and require a new 510(k) because
434 the different environments have different levels of professional
435 healthcare supervision or user training and offer different

⁵ A home use medical device is a medical device intended for users in any environment outside of a professional healthcare facility. This includes devices intended for use in both professional healthcare facilities and homes. See FDA’s Home Use Devices website for more information:
<http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/HomeHealthandConsumer/HomeUseDevices/default.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 436 environmental challenges, such as presence of other electronic
437 devices that can cause electromagnetic interference, different
438 levels of cleanliness, or shocks and vibrations associated with
439 patient travel or ambulatory use.
440
- 441 (2) Changes to enable use of the device by a different user.
- 442 • Similar to changes in use environment, how this type of change
443 affects a device’s risk profile depends on the difference in users.
444 Changes between similar types of users, such as changes between a
445 general physician and a specialist may not significantly affect a
446 device’s risk profile. Changes that enable unsupervised use by a
447 lay user as opposed to use by a healthcare provider (professional
448 use to lay use or prescription use to over-the-counter use),
449 however, are likely to significantly affect the device’s risk profile
450 and require a new 510(k) due to the different levels of user
451 training.
452
- 453 (3) Changes in the indications for use to a more specific use than the
454 currently cleared general indication.
- 455 • Manufacturers should carefully consider the potential effects on
456 their device’s risk profile in making these changes, as they are
457 among the most difficult to assess. If a change of this type has the
458 potential to expand device use to different users, different use
459 environments, use in or on a different type of joint, organ, bone,
460 vasculature, or tissue, use in different patient populations, or new
461 therapeutic or diagnostic uses, it should be evaluated using the
462 guidance provided above.
- 463 • FDA’s [*Guidance for Industry: General/Specific Intended Use*](#)
464 provides information on when a specific indication for use is
465 reasonably included within a general indication for use for
466 purposes of determining substantial equivalence, i.e., whether a
467 510(k) can be cleared or whether, instead, a PMA or de novo
468 submission is required. The factors discussed therein – particularly
469 those discussing the risk and public health impact of an indication
470 change – may be helpful to consider in deciding whether to submit
471 a new 510(k) for a change to an existing device, but that guidance
472 should not be used in and of itself to justify that a new 510(k) is
473 not required. The General/Specific guidance is not intended to
474 provide guidance on when a new 510(k) is required for changes to
475 an existing device.
476
- 477 (4) Changes in frequency or duration of use.
- 478 • Changes in the frequency or duration of use of a device include
479 changes indicating that a device can or should be used more or less
480 often, changes indicating that a device can perform a task or treat a

Contains Nonbinding Recommendations

Draft – Not for Implementation

481 condition in or for a different duration of time, or changes from
482 periodic to continuous monitoring. Manufacturers should evaluate
483 the effect such changes could have on the performance of a device,
484 and whether such changes significantly affect the device’s risk
485 profile.

- 486
487 (5) Changes concerning the compatibility or interoperability of a device
488 with other devices, components, or accessories.

489
490 An example of such changes would include changes indicating an IVD
491 reagent for use with a new system. To evaluate whether these changes
492 affect the device’s risk profile, manufacturers should carefully
493 consider the following factors:

- 494 • Differences between other devices, components, or accessories in
495 previously cleared indications and in the modified indications.
496 Manufacturers should be able to clearly identify and analyze the
497 risks associated with such differences, including whether the
498 change may affect biocompatibility, performance, connectivity,
499 etc. If the change is to indicate compatibility with a type of device,
500 component, or accessory that has not been indicated as compatible
501 previously, that change will likely require a new 510(k).
- 502 • The criticality of the other device, component, or accessory should
503 be factored in; the more critical the other device, component, or
504 accessory is to overall system function, the more likely a labeling
505 change regarding compatibility or interoperability could
506 significantly affect safety or effectiveness.
- 507 • The labeling of the other device, component, or accessory should
508 be considered. If the change is to indicate compatibility or
509 interoperability with another device that is also labeled for use with
510 the subject device or device type, it is less likely that the change
511 introduces a compatibility or interoperability issue that could
512 significantly affect safety or effectiveness.
- 513 • IVD manufacturers should see also FDA’s [Replacement Reagent](#)
514 [and Instrument Family Policy](#) guidance.

515
516 If the modification is a substantive change in the indications for use, a new 510(k) is
517 likely required. Otherwise, proceed to **A2**.

- 518
519 **A2. Does the change add or delete a contraindication?** Changes in the labeled
520 [contraindications](#) for device use generally could significantly affect safety or
521 effectiveness of a device and should typically be reviewed by the Agency, however,
522 FDA recognizes that, in general, the addition of a contraindication based on new
523 information is important to public health. Because of this, manufacturers are
524 encouraged to add new contraindications to their labeling and to notify existing users

Contains Nonbinding Recommendations

Draft – Not for Implementation

525 of their device as expeditiously as possible whenever a pressing public health need
526 arises. The new labeling should be submitted to FDA as part of a new 510(k) that is
527 prominently labeled “change being effected” (CBE, in Figure 2- Flowchart A). FDA
528 does not intend to take enforcement action against a device marketed with the
529 modified labeling that is submitted as part of a new CBE 510(k) while the 510(k) is
530 pending. Manufacturers should ensure they are thoroughly familiar with the definition
531 of a contraindication in such situations.

532
533 Deletion or modification of a contraindication usually requires the submission of a
534 new 510(k) prior to effecting the change, because this type of labeling change
535 typically substantively changes the indications for use. Deletions of contraindications
536 would expand the indications for use. For example, if a physical restraint was
537 contraindicated for use with individuals weighing less than 100 pounds because of
538 established life-threatening and other serious adverse events, and the manufacturer
539 subsequently wishes to remove this contraindication, a new 510(k) is likely required.

540
541 Similar to changes in indications for use, modifications that clarify or reword a
542 contraindication without affecting the substance of the contraindication would not
543 typically require a new 510(k).

544
545 If the change adds or deletes a contraindication, a new 510(k) is likely required.
546 Otherwise, proceed to **A3**.

547
548 **A3. Is it a change in warnings or precautions?** In order to facilitate a continuous
549 upgrading in device labeling, manufacturers should monitor device usage and
550 promptly revise the warnings and [precautions](#) section(s) based on user experience.
551 Events that precipitate changes of this type may be those reported under the medical
552 device reporting regulation (MDR), 21 CFR Part 803. New 510(k)s for such labeling
553 changes are generally not required, however, manufacturers should first proceed to
554 **A5.1** and **A5.2** and carefully consider whether the changes could affect the
555 indications for use or the device’s risk profile.

556
557 **A4. Does the change affect the instructions for use or other pieces of the labeling?**
558 Device labeling may be changed for a multitude of reasons. Many labeling changes
559 result from attempts to clarify labeling. Manufacturers should consider points **A5.1**
560 and **A5.2**, and if the change could affect how the device is used in practice. Labeling
561 changes that provide clarification without changing the meaning of the labeling
562 would generally not result in the need to submit a new 510(k).

563
564 **A5.1 Could the change affect the indications for use?** It is important to note that changes
565 to other parts of the labeling, such as the instructions for use, can affect the
566 indications for use even if the indications for use statement itself does not change.
567 Whether a labeling change can affect the indications for use will be device dependent.
568 As mentioned above, changes that could affect the indications for use of a device
569 generally require a new 510(k).

Contains Nonbinding Recommendations

Draft – Not for Implementation

570
571
572
573
574
575
576
577
578
579
580
581
582
583
584
585
586
587
588
589
590
591
592
593
594
595
596
597
598
599
600
601
602
603
604
605
606
607
608
609
610
611
612
613

Examples of labeling changes that could affect indications for use include:

- (1) Adding additional or new instructions on how to interpret diagnostic data from a diagnostic device.
- (2) Adding a new procedural technique not described in the original labeling.
- (3) Adding instructions for device use in a different patient population.
- (4) Adding instructions for device use in a different type of joint, organ, bone, vasculature, or tissue.
- (5) Changes from single-use to multiple use.

If the change affects the indications for use, a new 510(k) is likely required. Otherwise, proceed to **A5.2**.

A5.2 Does a risk assessment of the changed device identify any new risks or significantly modified existing risks? Changes to the labeling can also affect a device’s risk profile by affecting the way the device is used. As discussed in Question 1 of the Main Flowchart, if a change to labeling is intended to significantly affect safety or effectiveness by mitigating a new risk or an increased probability or severity of a known risk, that change likely requires a new 510(k), particularly if the new risk or increased risk has resulted in a recall, adverse events, or change in the acceptability of the risk. For labeling changes that are not intended to mitigate risks, but could affect a device’s risk profile, manufacturers should consult Section E and consider whether the change creates or significantly modifies risks. As part of that evaluation, manufacturers should consider whether changes to labeling could introduce human factors or usability issues that could significantly affect users’ understanding of the labeling and use of the device. Changes that significantly affect a device’s risk profile likely require a new 510(k).

Examples of labeling changes that may affect the device’s risk profile include:

- (1) Use of a product for a duration/frequency that is different than what is described in the labeling of the cleared device.
- (2) Changing from labeling a device as non-sterile to labeling it as sterile.
- (3) Changes concerning device compatibility or interoperability with other devices, components, or accessories. See **A1**, above. Manufacturers should consider the factors discussed there to determine whether these changes will require a new 510(k).

If the change significantly affects the device’s risk profile, a new 510(k) is likely required. Otherwise, a 510(k) is likely unnecessary for a labeling change, unless otherwise indicated above.

Contains Nonbinding Recommendations

Draft – Not for Implementation

614 FDA believes that, if manufacturers follow this approach to changes in device labeling, only
615 necessary new 510(k)s (those changes that could significantly impact safety and
616 effectiveness) will be submitted, while the submission of unnecessary new 510(k)s (those
617 that could not significantly affect safety and effectiveness) will be minimized. At the same
618 time, manufacturers should be able to retain the flexibility to improve their labeling to assure
619 safe and effective use of their devices.
620

621 **B. Technology, Engineering, and Performance Changes**
622

623 These types of changes encompass a broad span of design activities, from minor engineering
624 changes in a circuit board layout to a change from electromechanical to microprocessor
625 control of device function. Flowchart B illustrates the decision-making logic scheme for such
626 technology, engineering or performance changes to a device. All changes should be
627 evaluated using this scheme, and then the changes should be verified and/or validated
628 according to the QS requirements, 21 CFR 820.30(i). If the results of the verification and/or
629 validation raise any unexpected issues, the decision of whether a new 510(k) is required
630 should be re-evaluated per B5.4.

DRAFT

Contains Nonbinding Recommendations

Draft – Not for Implementation




Figure 3 - Flowchart B

631
632
633
634
635
636

B1. Is the device an *in vitro* diagnostic device? If the device is an IVD, refer to the later section of this guidance which is specific to technology, engineering, and

Contains Nonbinding Recommendations

Draft – Not for Implementation

637 performance changes in IVDs (Section D – Technology, Engineering, Performance,
638 and Materials Changes for In Vitro Diagnostic Devices).

639

640 **B2. Is it a control mechanism, operating principle, or energy type change?**

641

642 **Control mechanism changes:** A [control mechanism](#), for the purpose of this
643 guidance, is the manner by which the actions of a device are directed. Almost all
644 changes in the control mechanism for a device could significantly affect safety and
645 effectiveness. Therefore, such changes will normally require the submission of a new
646 510(k). This is also true for changes in operating principle as well as for changes in
647 energy type (discussed below). Changes of these types tend to be more revolutionary
648 than evolutionary.

649

650 One example of a control mechanism change would be a change from analog to
651 digital control of a medical device. While the change to digital control can markedly
652 improve device [performance specifications](#) and effectiveness, the integration of a
653 digital control into a previously all-analog system is complex and usually undertaken
654 only as part of a major redesign of a product. Thus, it would be rare that a new 510(k)
655 would not be required. Most often, such changes in control mechanism represent the
656 introduction of a new product line.

657

658 Other changes in control mechanism of a similar nature would also likely require a
659 new 510(k). An example of such a change would be the change from pneumatic to
660 electronic control of a respiratory care device.

661

662 **Operating principle changes:** Similar to a control mechanism change, a change in
663 [operating principle](#) would also usually require the submission of a new 510(k). An
664 example of a new operating principle for a device would be changing the image
665 reconstruction algorithm used in a computed tomography x-ray system from simple
666 back projection to a new, more radiation-efficient method. In this case, testing both at
667 the bench and in the clinic would be necessary to support a finding of substantial
668 equivalence for the new device. Another example would be a change in a water
669 droplet dispersal method used by a respiratory gas humidifier from piezoelectric
670 material to a wick and fan method. The two mechanisms use the same design
671 principle, but apply it in different ways. The differences between the two could
672 significantly affect safety and effectiveness.

673

674 Such changes may also be accompanied by significant labeling changes and,
675 sometimes, by a need for operator retraining to ensure continued safe and effective
676 operation.

677

678 **Energy type changes:** The submission of a new 510(k) will usually be required for
679 [energy type](#) changes. These changes include both energy output and input changes. A
680 change from emitting microwave energy to radiofrequency (RF) energy would be an
681 example of an energy output change; this type of change would likely be part of a

Contains Nonbinding Recommendations

Draft – Not for Implementation

682 significant redesign. An example of an energy type input change is a modification
683 from AC to battery power; this type of change is usually part of a redesign to provide
684 a portable device that can be used under different environmental conditions than the
685 original device. Such a change would normally be accompanied by significant
686 labeling changes, including a new or expanded indication for use. Note that this type
687 of change does not include a change in voltage, such as from 3V to 9V operation or a
688 change between different types of batteries, such as from NiCad to lead acid storage
689 batteries. Such changes should be considered changes in performance specifications
690 or device design, as discussed at decision point **B5**.

691

692 **B3. Is it a change in sterilization, cleaning, or disinfection?** Changes in sterilization,
693 cleaning, or disinfection should be carefully assessed. If there is a change of this type,
694 proceed to **B3.1**.

695

696 **B3.1 Is it a change to an “established category B” or “novel” sterilization method,
697 does the change lower the sterility assurance level, or is it a change to how the
698 device is provided?** Changes from “established category A” sterilization methods to
699 “established category B” or “novel” sterilization methods generally require a new
700 510(k). Changes from one “established category A” method to another “established
701 category A” method, or from an “established category B” or “novel” method to an
702 “established category A” method, should be evaluated under **B3.2**. See FDA’s
703 guidance [Submission and Review of Sterility Information in Premarket Notification
704 \(510\(k\)\) Submissions for Devices Labeled as Sterile](#) for a discussion of sterilization
705 methods.

706

707 If the [sterility assurance level](#) (SAL) is lowered, manufacturers should consider
708 whether device safety or effectiveness may be compromised by the new level. In
709 general, reductions in SAL require new 510(k) submissions unless the SAL remains
710 better than 10^{-6} . Note that changes to cleaning and disinfection processes for
711 reprocessed devices can also affect the bioburden levels on a device, which may
712 invalidate subsequent processing steps such as sterilization; manufacturers should
713 carefully consider whether these changes could significantly affect the safety or
714 effectiveness of the device. It is likely that changes to [reprocessing](#) procedures for
715 devices listed in Appendix E of FDA’s guidance, [Reprocessing Medical Devices in
716 Health Care Settings: Validation Methods and Labeling](#), could significantly affect
717 safety or effectiveness. FDA has identified the devices there as a subset of medical
718 devices that pose a greater likelihood of microbial transmission and represent a high
719 risk of infection (subclinical or clinical) if they are not adequately reprocessed.

720

721 Some changes to how a device is provided to the user or patient could also
722 significantly affect safety or effectiveness. For the purposes of this question, “how a
723 device is provided” refers to whether the device is provided sterile or non-sterile, and
724 to whether the device is provided for (1) single-patient, single-use, (2) single-patient,
725 multi-use, or (3) multi-patient, multi-use. If a device is changed from (1) to (2), (1) to
726 (3), or (2) to (3), i.e., provided for more patients and/or more uses, a 510(k) is likely

Contains Nonbinding Recommendations

Draft – Not for Implementation

727 required. However, the reverse would not be true; it would be unlikely that a change
728 from (3) to (2), (3) to (1), or (2) to (1) could significantly affect safety or
729 effectiveness and therefore would not likely require a new 510(k). In addition, if a
730 device that was originally provided sterile is modified to be provided non-sterile –
731 either to be sterilized by the user or to be used without sterilization – a new 510(k) is
732 likely required. A new 510(k) is also likely required if a device originally provided
733 non-sterile is modified to be provided sterile.

734

735 If the answer to this question is yes, a new 510(k) is likely required. If the answer is
736 no, proceed to **B3.2**.

737

738 **B3.2 Could the change significantly affect the performance or biocompatibility of the**
739 **device?** Changes in the [method of sterilization](#), cleaning, or disinfection have the
740 potential to change material or performance characteristics of a device. This is
741 particularly true of the properties of polymeric materials or surface coatings. When
742 manufacturers make changes in sterilization, cleaning, or disinfection methods, they
743 should consider whether the properties or specifications of the device could be
744 significantly affected.

745

746 To determine whether the sterilization, cleaning, or disinfection change could
747 significantly affect device performance, the manufacturer should consider known
748 information on the sterilization, cleaning or disinfection method, its parameters, and
749 the material being sterilized, cleaned, or disinfected, and determine if there are any
750 new or significantly modified risks associated with using the proposed method and its
751 parameters with the device's materials of construction. If there are new or
752 significantly modified risks (see Section E), this likely indicates that the change could
753 significantly affect the device's safety or effectiveness. Note also that if verification
754 and/or validation of the new methods show any unexpected results, manufacturers
755 should re-evaluate whether a new 510(k) is required (see **B5.4**).

756

757 Sterilization, cleaning, or disinfection changes may also affect the biocompatibility of
758 a device. For instance, changes to an ethylene oxide sterilization process may leave
759 increased ethylene oxide residuals on the device surface, or changes to a cleaning
760 process may incorporate chemicals that are inappropriate for use with a patient-
761 contacting device. Manufacturers should consider whether sterilization, cleaning, or
762 disinfection changes could significantly affect the biocompatibility of their device. If
763 a manufacturer determines their cleaning, disinfection, or sterilization change could
764 significantly affect the performance or biocompatibility of the device, a new 510(k) is
765 likely required. Otherwise, it is unlikely a 510(k) is required as a result of this type of
766 change.

767

768 **B4. Is there a change in packaging or expiration dating?** If yes, proceed to **B4.1**.

769

770 **B4.1 Is the same method or protocol, as described in a previously cleared 510(k), used**
771 **to support the change?** Generally, changes in device [packaging](#) or changes in the

Contains Nonbinding Recommendations

Draft – Not for Implementation

772 [expiration date](#) for use of a device do not require a new 510(k). FDA relies on the QS
773 regulation (21 CFR Part 820) to reasonably assure the safety and effectiveness of
774 devices with these types of changes. This is true whether or not the manufacturer
775 applies an expiration date because of package integrity considerations, e.g., sterility,
776 or because of a finite shelf-life of the device. However, where methods or protocols
777 that are not described in a previously cleared 510(k) are used to support new package
778 integrity or shelf-life claims, a new 510(k) is likely required.
779

780 **B5. Is it any other change in design (e.g., dimensions, performance specifications,**
781 **wireless communication, components or accessories, or the patient/user**
782 **interface)?** These types of design or engineering changes encompass everything from
783 the routine specification changes necessary to maintain or improve device
784 performance as a result of feedback from users, field or plant personnel, etc., up to
785 and including significant product redesign. The bullets below highlight some, but not
786 all, of these changes, and provide points to consider for each type of change.
787

788 • **Dimension changes:** In determining whether a new 510(k) is required for these
789 types of changes, per **B5.1-B5.4**, the manufacturer should consider not only the
790 size of the dimension or [dimensional specification](#) change, but the criticality of
791 the modified dimension. The more critical the dimensions being modified are to
792 the safe and effective operation of the device, the more likely it is that the change
793 could significantly affect safety or effectiveness. For instance, a 1 mm change to
794 the diameter of a working channel of an endoscope is more likely to significantly
795 affect safety or effectiveness than a 1 mm change to the length of an endoscope.
796

797 If a modified dimension is within a range of dimensions previously cleared for the
798 original device, a new 510(k) would not typically be required. For instance, if the
799 original device was cleared with two models that were 2 and 4 mm in diameter,
800 and the modified device of the same length has a diameter of 3 mm, a new 510(k)
801 is likely not required for this change.
802

803 • **Device performance changes:** This category covers a broad range of changes. As
804 discussed in the Main Flowchart, Question 1, changes that are intended to
805 significantly affect device safety or effectiveness likely require a new 510(k).
806 Changes that are not intended to affect device safety or effectiveness should be
807 considered per **B5.1-B5.4**.
808

809 • **Wireless communication changes:** Changes to device communication between
810 device components or between the modified device and other products,
811 particularly from wired to wireless, may change a device's risk profile by
812 introducing or modifying risks regarding data transmission or cybersecurity.⁶

⁶ See FDA's webpage on cybersecurity in medical devices,
<http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/ConnectedHealth/ucm373213.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

813 Changes to employ wireless communication in devices where it was previously
814 not used are likely to significantly affect safety or effectiveness and likely require
815 a new 510(k). This is particularly true when wireless communication is used to
816 control device operations. When evaluating other changes, including a change to a
817 different wireless communication protocol, the factors in **B5.1-B5.4** should be
818 taken into account in determining whether a new 510(k) is required.
819

- 820 • **Components or accessories:** Changes to components or accessories could, in
821 some cases, significantly affect the safety or effectiveness of the device as a
822 whole. In **B5.1**, manufacturers should consider whether changes to the device or
823 any of its components or accessories affect the use of other components or
824 accessories, or if changes to a component or accessory could lead a device to be
825 used in a new way. In **B5.2**, manufacturers should consider whether changes to
826 the device or any of its components or accessories could disrupt compatibility
827 between the device and its components or its accessories, and whether these
828 changes could lead to a significant change in the device’s risk profile.
829
- 830 • **Changes in the [human factors of the patient or user interface](#):** A device [user](#)
831 [interface](#) includes all points of interaction between the product and the user,
832 including elements such as displays, controls, and packaging. User interface
833 changes refer to changes in the way in which a patient or user interacts with a
834 device, including, for example, the way in which the device presents alarms to the
835 user, the layout of the control panel, the mode of presentation of information to
836 the user or patient, and the way in which the device physically interacts with the
837 user and/or patient (e.g., the way in which a CPAP mask attaches to a patient’s
838 face, or the way a surgical instrument is designed to fit in a surgeon’s hand). Note
839 that this type of change includes changes that modify a user workflow (tasks
840 performed by a user in order to complete their work). Manufacturers should
841 consider the risk impact of changes in user workflow; for example, providing new
842 information to the user or modifying the manner in which information is
843 presented may impact user comprehension. In addition, changing the layout of
844 device controls may impact device use differently in different use scenarios. For
845 more information on applying human factors in medical devices, see FDA’s
846 guidance [Applying Human Factors and Usability Engineering to Optimize](#)
847 [Medical Device Design](#).
848

849 Changes intended only to increase user or patient comfort when interacting with
850 the device may be particularly difficult to evaluate. These changes will typically
851 not require a new 510(k), but some changes made for the comfort of the user or
852 patient could significantly affect safety or effectiveness. For example, if a surgical
853 handpiece is redesigned to move a motor closer to the surgeon’s hand or the
854 surgical site, any heating of the motor will be more likely to affect the surgeon or
855 patient and could result in burns. Manufacturers should evaluate changes to a user
856 interface and whether they significantly affect safety or effectiveness in answering
857 **B5.1-B5.4**.

Contains Nonbinding Recommendations

Draft – Not for Implementation

858
859
860
861
862
863
864
865
866
867
868
869
870
871
872
873
874
875
876
877
878
879
880
881
882
883
884
885
886
887
888
889
890
891
892
893
894
895
896
897
898
899
900

Changes in design should be considered, along with the above bulleted points, in answering **B5.1-B5.4**.

B5.1 Does the change significantly affect the use of the device? As with a labeling change, if a design change significantly affects how a device may be used, a new 510(k) is likely required. Manufacturers should consider whether the change increases the likelihood that the device will be used by a broader or different group of users who have less training regarding safe and effective use of the device (e.g., lay users instead of clinicians, or general practitioners instead of surgeons) and whether that change affects the risk profile of the device. If the change significantly affects the risk profile (see Section E), a new 510(k) is likely required.

Manufacturers should also consider whether the change increases the likelihood that the device will be used in a new environment, and whether the new environment affects the risk profile of the device. If the change facilitates use in a completely different environment (e.g., from hospital to home use, or from hospital to ambulance transport), this typically will introduce new or significantly modified risks and will likely require a new 510(k). If the change facilitates use only in similar environments, the risk profile of a device may also be changed, but this is less likely to require a new 510(k). In deciding whether a change that allows use of the device in a new environment could significantly affect the safety or effectiveness of the device, manufacturers should consider differences in environmental specifications such as:

- (1) Temperatures and humidity that might affect device operation;
- (2) Noises that might drown out the sound of auditory alarms;
- (3) Exposure to water, soils, or light that might affect device operation;
- (4) Presence of other devices or equipment that may cause electromagnetic interference; and
- (5) Possible use in magnetic resonance imaging (MRI).

If the change introduces new or significantly modified risks, a new 510(k) is likely required.

If the change significantly affects use of the device, a new 510(k) is likely required. If it does not, proceed to **B5.2**.

B5.2 Does a risk assessment of the changed device identify any new risks or significantly modified existing risks? As discussed in the Guiding Principles and Section E, the manufacturer should conduct a risk assessment for any modified device. New risks, changes to the acceptability of previously identified risks, or changes to device features that may be critical to the device’s safe or effective operation will likely require new 510(k)s.

Contains Nonbinding Recommendations

Draft – Not for Implementation

901 Manufacturers should carefully consider whether changing one aspect or feature of a
902 device’s design might affect a seemingly unrelated aspect or feature. For instance, a
903 dimensional or component change may affect the ability to reprocess a device or the
904 ability to regulate the temperature of an electronic device. Manufacturers should
905 evaluate these impacts of the change as part of their risk assessment.

906

907 If a risk assessment does not identify any new risks or significantly modified existing
908 risks or effectiveness issues per Section E, proceed to **B5.3**.

909

910 **B5.3 Are clinical data necessary to evaluate safety or effectiveness for purposes of**
911 **design validation?** Whenever a manufacturer recognizes that clinical data are needed
912 because bench testing or simulations are not sufficient to assess the impact of the
913 change on safety or effectiveness to validate the design change, a new 510(k) is likely
914 required. For the purposes of this question, clinical data does not include data not
915 used for design validation, such as user or patient preference testing.

916

917 If clinical data are unnecessary to evaluate safety and effectiveness for purposes of
918 design validation, proceed to **B5.4**.

919

920 **B5.4 Do design verification and/or validation activities produce any unexpected issues**
921 **of safety or effectiveness?** All changes to device design should undergo some level
922 of design verification and/or validation or evaluation to ensure that the device
923 continues to perform as intended. See 21 CFR 820.30. As discussed in the Guiding
924 Principles, manufacturers should make an initial risk-based assessment of whether a
925 change requires a new 510(k). If the manufacturer determines after an initial
926 assessment that a new 510(k) is not required, the manufacturer should conduct routine
927 verification and validation activities to ensure that no new issues of safety or
928 effectiveness are raised. If successful application of routine verification and
929 validation activities confirms the initial assessment, manufacturers should proceed
930 with the design change and document their assessment.

931

932 Occasionally, routine verification and validation activities may either produce
933 unexpected results or otherwise prove to be inadequate to verify and/or validate the
934 modified design. In such instances, the manufacturer likely is required to submit a
935 new 510(k).

936

937 If a manufacturer encounters unexpected results performing routine verification and
938 validation activities – for example, the device does not perform as expected, pre-
939 specified acceptance criteria are not met, or testing demonstrates unexpected safety or
940 effectiveness issues – the manufacturer should analyze the results carefully. The
941 initial risk assessment should be re-evaluated, and if changes to that assessment are
942 necessary, the manufacturer should re-evaluate whether the device change could
943 significantly affect safety or effectiveness. If different verification and/or validation
944 test methods or acceptance criteria are necessary to produce the expected results, it is

Contains Nonbinding Recommendations

Draft – Not for Implementation

945 likely that the change could significantly affect safety or effectiveness and thus a new
946 510(k) is likely required.

947
948 If the manufacturer determines prior to conducting verification and validation
949 activities that routine verification and validation activities are insufficient and the
950 design change necessitates a different verification and/or validation scheme or new
951 acceptance criteria, a new 510(k) is likely required. This does not mean that
952 manufacturers should not update test methods and acceptance criteria for verification
953 and validation activities in accordance with advances in science or relevant voluntary
954 consensus standards, but if the design change drives the need for a new testing
955 scheme or acceptance criteria (as opposed to advances in science or standards), it is
956 likely that the design change could significantly affect safety or effectiveness.

957
958 If the initial assessment determines a new 510(k) is not required, and verification and
959 validation activities are substantially unchanged (i.e., use the same test methods and
960 same acceptance criteria) and successful, then proceed to Section C.

961
962 For example, in order to better accommodate connection of a urinary drainage (Foley)
963 catheter to a collection apparatus, the manufacturer increases the length of the
964 catheter by several millimeters. The new length is outside of previously cleared
965 lengths for this device, however, the length change is not far outside cleared lengths.
966 Based on its risk assessment, the manufacturer does not expect the length change will
967 create any new risks or significantly affect existing risks. The manufacturer therefore
968 determines that the length change could not significantly affect the device's safety or
969 effectiveness, and does not require a new 510(k). The manufacturer subsequently
970 conducts design control activities, and verifies that the catheter functions safely and
971 effectively, as predicted, with no unexpected results. The manufacturer documents
972 these efforts and proceeds to production.

973
974 On the other hand, a manufacturer of monitoring devices wants to use a more
975 sensitive comparator circuit, and makes other design changes to accommodate the
976 more sensitive component. The design change is similarly evaluated in an initial risk
977 assessment based on models, calculations, etc., and a decision is made that the change
978 could not significantly affect the device's safety or effectiveness, and therefore the
979 changes do not require a new 510(k). However, as part of routine verification and
980 validation activities, tests with a simulator produce unexpected results, and additional
981 work is necessary to understand how and why this outcome occurred. The
982 manufacturer should carefully assess these results and whether new issues of safety or
983 effectiveness have been uncovered.

984


985 **C. Materials Changes**

986

Contains Nonbinding Recommendations

Draft – Not for Implementation

987 Firms making changes to the materials from which their device is manufactured should also
988 consider the other types of changes discussed above and their impact on the decision
989 regarding a new 510(k). For example, a material change, as discussed below, might also lead
990 to a change in the labeling of the device (e.g., the removal of a contraindication or the
991 addition of a new warning), or a change in specifications (e.g., a reduction in the strength of
992 the device). These collateral changes should be considered in addition to the logic scheme
993 described in this section.



994
995
996
997

Figure 4 - Flowchart C

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 998 **C1. Is the device an *in vitro* diagnostic device?** If the device is an IVD, refer to the later
999 section of this guidance which is specific to materials changes in IVDs (Section D –
1000 Technology, Engineering, Performance, and Materials Changes for In Vitro
1001 Diagnostic Devices).
1002
- 1003 **C2. Is this a change in material type, material formulation, chemical composition, or**
1004 **the material’s processing?** If there is any change in [material type](#), [formulation](#), or
1005 chemical composition, the answer to this question should be yes. Additionally, if
1006 there is any change in supplier or manufacturer material processing or finishing steps,
1007 the answer should also be yes. The biocompatibility and physical properties of a
1008 finished device depend not only on the materials, but also on the processing of the
1009 materials, manufacturing methods (including the sterilization process), and the
1010 manufacturing residuals that may be present on the finished device. Changes of this
1011 type should be further evaluated for their potential impact on safety and effectiveness.
1012 The subsequent questions, such as **C4** and **C4.1**, address whether the change is
1013 significant.
1014
- 1015 Many material changes result from [material supplier](#) changes, including changes
1016 made by a material supplier, or changes from one supplier to another. When these
1017 types of changes occur, the manufacturer should utilize their quality system process
1018 to analyze the material and determine the extent of the change made, as this analysis
1019 might impact answers to subsequent questions.
1020
- 1021 If there is a change in material type, material formulation, chemical composition, or
1022 the material’s processing as described above, proceed to **C3**. Otherwise it is unlikely
1023 a new 510(k) is required as a result of a materials change.
1024
- 1025 **C3. Will the changed material directly or indirectly contact body tissues or fluids?**
1026 Both direct and indirect patient and user contact should be considered in answering
1027 this question. Direct contact is when a material touches any tissue or bodily substance
1028 of a patient or user while the material is still in or on the patient or user. Indirect
1029 contact is when a material has the potential to come into contact with any tissue or
1030 bodily substance through some intervening material (such as a liquid or gas) by first
1031 coming in contact with the intervening material, which subsequently comes in contact
1032 with the patient tissue or bodily substance. For example, materials in a catheter hub
1033 (the part of the catheter which is external to the patient) can contact the patient
1034 indirectly if fluids or drugs are infused through the hub and directly into the patient.
1035
- 1036 While most implant materials contact patients, there are some exceptions. For
1037 example, the internal contents of spinal cord stimulators are not patient-contacting;
1038 they are hermetically sealed so that there is no material transfer, fluid transfer, or
1039 leeching out of any material internal to the device.
1040
- 1041 If the changed material directly or indirectly contacts body tissues or fluids, proceed
1042 to **C4**. If the changed material does not contact body tissues or fluids, proceed to **C5**.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1043
1044
1045
1046
1047
1048
1049
1050
1051
1052
1053
1054
1055
1056
1057
1058
1059
1060
1061
1062
1063
1064
1065
1066
1067
1068
1069
1070
1071
1072
1073
1074
1075
1076
1077
1078
1079
1080
1081
1082
1083
1084
1085
1086

C4. Does a risk assessment identify any new or increased biocompatibility concerns? Manufacturers should conduct a risk assessment, which may include an assessment of the device’s toxicological and physical properties, of any changed materials that may contact the patient or user to determine if there are any new or increased biocompatibility concerns. An example of a new concern would be a material change that requires a new type of biocompatibility test, such as an implantation test, that was not required for the original device. An example of an increased concern would be where a new chemical component added to a material requires a genotoxicity analysis of that component (because, for instance, the particular component is noted in the literature as potentially genotoxic), but the original device already required a genotoxicity analysis. ISO 10993-1, *Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing Within a Risk Management Process* may be useful in this assessment.

The answer to **C4** may be no if a knowledgeable individual reviews the differences in chemical composition or physical properties and determines that the change is minor enough that there is no new concern about biocompatibility. See FDA’s [*Use of International Standard ISO-10993, “Biological Evaluation of Medical Devices Part 1: Evaluation and Testing,”*](#) for further information on how to review such differences.

A supporting toxicological assessment can be based on an analysis of the chemical formulations or the results of chemical characterization tests if the detailed formulation is not available (i.e., when the material is provided by a supplier and the formulation is proprietary). If, however, this analysis identifies new chemical entities or other properties that are either novel or have the potential to generate adverse biocompatibility responses, such as genotoxicity, a new 510(k) may be required.

If a risk assessment identifies any new or increased biocompatibility risks, consider the questions in **C4.1**. If no new or increased biocompatibility risks are identified, proceed to **C5**.

C4.1 Has the manufacturer used the same material in a similar legally marketed device? Manufacturers who have identified possible biocompatibility concerns in their risk assessment (**C4**) should consider whether they have used the same material, in its final, finished state, in another one of its own legally marketed devices that has been cleared or approved by the FDA. If the manufacturer has used the same material in a similar device that has been cleared or approved by the FDA (this would typically involve a biocompatibility evaluation), and there is no postmarket evidence of biocompatibility issues with the device, that may provide evidence that the material will be biocompatible in its new application in the changed device as well and the manufacturer can answer yes to this question.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1087 It is important to note that in order to answer yes to this question, the material in
1088 question should have the same formulation or chemical composition and be subjected
1089 to the same processing, including sterilization (i.e., the comparison should be between
1090 materials as they are applied in the final finished device, not between raw materials).
1091 In addition, the size and geometry of the changed device or component should not
1092 affect the curing of the polymer or result in more material in the new device or
1093 component.

1094
1095 The previously cleared or approved device should have the same or a more risky type
1096 of contact and the same or a longer duration of contact. For example, if a
1097 manufacturer intends to use a new material in a limited exposure application (<24
1098 hours), and the manufacturer has used that same material in a cleared or approved
1099 device for prolonged exposure (24 hours to 30 days), then it is unlikely that a new
1100 510(k) will be required for this change. If the modified device is intended to have a
1101 riskier category of contact (e.g., mucosal membrane contact is riskier than contact
1102 with intact skin, and blood contact is riskier than tissue/bone contact) or a longer
1103 duration of contact, then the manufacturer should answer no to this question. Contact
1104 may be either direct or indirect.

1105
1106 Manufacturers should not compare their changed material to materials in other
1107 manufacturers' legally marketed devices, unless the exact formulation and processing
1108 of the device, which may affect the safety and effectiveness of the final finished
1109 product, can be verified.

1110
1111 If the manufacturer has used the same material in a similar legally marketed device,
1112 proceed to **C5** to determine if the material change could affect device performance. If
1113 the manufacturer has not used the same material in a similar legally marketed device,
1114 a new 510(k) is likely required.

1115
1116 **C5. Could the change affect the device's performance specifications?** Manufacturers
1117 should consider whether the material change could affect the performance of the
1118 device by affecting its strength, hardness, etc. Manufacturers should also consider
1119 whether the new material could be affected by any labeled cleaning, disinfection, or
1120 sterilization process of the device. If the answer to this question is yes, manufacturers
1121 should proceed to **B5** above and consider whether the change could significantly
1122 affect the safety or effectiveness of the device. If the change could not affect the
1123 device's performance specifications, it is unlikely the change could significantly
1124 affect safety or effectiveness, and the manufacturer should document the change.

1125

1126 **D. Technology, Engineering, Performance, and**
1127 **Materials Changes for In Vitro Diagnostic Devices**

1128

1129 Firms making technology, engineering, performance, or materials changes to their IVD
 1130 should also consider the other types of changes discussed above in Section A, Labeling
 1131 Changes, and their impact on the decision regarding submission of a new 510(k). For
 1132 example, a material change, as discussed below, might also be considered a design change
 1133 and/or might engender a change in the labeling of a device (e.g., the removal of a
 1134 contraindication, addition of a new warning, or a change in the measuring range). These
 1135 collateral changes should be considered also when applying the logic scheme described in
 1136 this section.
 1137




Figure 5 - Flowchart D

1138
 1139
 1140
 1141
 1142
 1143
 1144
 1145

D1. Does the change alter the operating principle of the IVD?

In most cases, a technology, engineering, performance, or material change that alters the operating principle of an IVD could significantly affect safety and effectiveness, in which case a new 510(k) is required (21 CFR 807.81(a)(3)(i)). Changes in

Contains Nonbinding Recommendations

Draft – Not for Implementation

1146 technology, engineering, performance, or materials of an IVD can include changes
1147 made to reagents or changes to a test method or protocol, among other things.

1148

1149 Examples of changes in technology, engineering, performance, or materials that
1150 likely alter the operating principle of the IVD and for which a new 510(k) is likely
1151 required include: changes from liquid to solid reagent; changes from
1152 radioimmunoassays (RIA) to non-RIAs; changes in the antibody; changes in detection
1153 reagents; changes in critical reaction components; changes in conjugates. Examples
1154 of changes in technology, engineering, performance, or materials that might alter the
1155 operating principle of the IVD include changes in calibration materials and quality
1156 control materials; changes in substrates; changes in specimen type changes in
1157 specimen processing; changes in incubation times and temperatures. A new 510(k) is
1158 not necessarily required for all changes in technology, engineering, performance, or
1159 materials for IVDs that alter the operating principle of an IVD. However, when, for
1160 example, such changes introduce novel technology that could have an impact on the
1161 ability of the device to extract, isolate, or detect the analyte(s) and could therefore
1162 affect the value assigned to the specimen, or could produce deviations in device
1163 performance that would result in modified reporting of performance in labeling, then
1164 a new 510(k) is likely required.

1165

1166 Examples of changes in technology, engineering, performance, or materials of an
1167 IVD which do not ordinarily affect the operating principle include: changes to
1168 external packaging, changes to use a new lot or batch for the same antibody or
1169 enzyme, changes to a new vendor for the same reagent, and changes in concentrations
1170 of packaged reagents provided the same diluted concentration was used in the assay.

1171

1172 If such a change to an IVD does not alter the operating principle of the IVD, proceed
1173 to D2.

1174

1175 **D2. Is the change identified in a device-specific guidance or classification regulation?**

1176

1177 In the case of some IVDs, FDA has published device-specific guidance documents,
1178 which provides resources to manufacturers on specific issues related to those devices.
1179 A searchable listing of these device-specific guidances can be found on [FDA's](#)
1180 [website](#). When a device-specific guidance identifies a change that FDA has
1181 determined could significantly affect safety or effectiveness, a new 510(k) is
1182 generally required under 21 CFR 807.81(a)(3)(i). Additionally, in the case of some
1183 IVDs, FDA has established specific requirements (e.g., special controls) that are
1184 identified in the classification regulation. If a classification regulation identifies a
1185 change that could significantly affect safety or effectiveness, a new 510(k) is
1186 required. Where a change is not identified in a device-specific guidance or
1187 classification regulation, proceed to D3.

1188

1189 **D3. Does a risk assessment of the changed device identify any new risks or**
1190 **significantly modified existing risks?**

Contains Nonbinding Recommendations

Draft – Not for Implementation

1191
1192
1193
1194
1195
1196
1197
1198
1199
1200
1201
1202
1203
1204
1205
1206
1207
1208
1209
1210
1211
1212
1213
1214
1215
1216
1217
1218
1219
1220
1221
1222
1223
1224
1225
1226
1227
1228
1229
1230
1231
1232
1233
1234
1235

As discussed in the Guiding Principles and Section E, the manufacturer of an IVD should conduct a risk assessment for any modified device. Changes in the technology, engineering, design, or material used in an IVD can affect the performance, including the analytical or clinical performance, of the device. Further, certain changes in an IVD could also present new or significantly modified risks apart from performance. These changes could affect the risk profile of the IVD, apart from the performance (e.g., transmission of pathogenic diseases, biocompatibility or sterility issues).

For IVDs, a manufacturer’s risk assessment identifies new risks or significantly modified existing risks when the risk assessment (1) indicates that the performance of the modified test could significantly change from the previously cleared performance claims or (2) identifies new risks or significantly modified existing risks, apart from performance. If a change could affect the analytical performance of a device, particular attention should be paid to the effect on device performance around the clinical decision point(s) (i.e., cut-offs, cut-points). When new risks or significantly modified existing risks have been identified, in general, the change to the IVD could significantly affect safety or effectiveness of the device and a new 510(k) is likely required. This includes a change that is clinically significant in terms of clinical decision making.

Changes to components or accessories could, in some cases, significantly affect the safety or effectiveness of an IVD as a whole. Manufacturers should consider in their initial risk assessment whether changes to the IVD or any of its components or accessories affect the use of other components or accessories, or if changes to a component or accessory could lead an IVD to be used in a new way. Manufacturers should also consider whether changes to the IVD or any of its components or accessories could disrupt compatibility between the device, its components, and/or its accessories, or whether these changes could significantly affect performance or the device’s risk profile.

Changes in the human factors of a patient or user interface could, in some cases, significantly affect the safety or effectiveness of an IVD as a whole. Manufacturers should evaluate in their initial risk assessment whether a change in the human factors of a patient or user interface could significantly change the performance of the IVD or presents new risks or significantly modified existing risks. A device user interface includes all points of interaction between the product and the user, including elements such as displays, controls, and packaging. User interface changes refer to changes in the way in which a patient or user interacts with a device, including, for example, the way in which the device presents alarms to the user, the layout of the control panel, the mode of presentation of information to the user or patient, and the way in which the device physically interacts with the user and/or patient. Note that these changes include those that modify a user workflow (tasks performed by a user in order to complete their workflow). Manufacturers should consider the risk impact of changes

Contains Nonbinding Recommendations

Draft – Not for Implementation

1236 in user workflow, such as providing new information to the user or modifying the
1237 manner in which information is presented, which may impact comprehension, or
1238 changing the layout of device controls, which may impact device use differently in
1239 different use scenarios. For more information on applying human factors in medical
1240 devices, see FDA’s guidance, [Applying Human Factors and Usability Engineering to](#)
1241 [Optimize Medical Device Design](#).

1242
1243 Changes intended only to increase user or patient comfort when interacting with the
1244 device may be particularly difficult to evaluate. These changes will typically not
1245 present new risks or modified existing risks, but some changes made for the comfort
1246 of the user or patient could significantly affect safety or effectiveness. Manufacturers
1247 should evaluate the potential of changes to a user interface as to whether they could
1248 significantly affect safety or effectiveness.

1249
1250 If a risk assessment indicates that that the performance of the modified IVD could not
1251 significantly change from the previously cleared performance claims, or that the
1252 modified IVD does not present new or significantly modified existing risks apart from
1253 performance, proceed to D4.

1254
1255 **D4. Do design verification and/or validation activities produce any unexpected issues**
1256 **of safety or effectiveness?**

1257
1258 As discussed above in the Guiding Principles, manufacturers should conduct an initial
1259 risk assessment of whether a change requires a new 510(k); if the initial decision
1260 following the risk assessment is that a new 510(k) is not required, the manufacturer
1261 should conduct design verification and/or validation activities to confirm the decision.
1262 Results of the design verification and/or validation activities may serve to aid a
1263 manufacturer in determining whether a technology, engineering, performance, or
1264 material change could significantly affect safety and effectiveness.

1265
1266 Generally, FDA’s 510(k) clearances of IVDs include specified performance claims or
1267 performance specifications. For IVDs, a new 510(k) is likely not required where (1)
1268 standard methods and established and justified criteria (e.g., clinically appropriate
1269 criteria or criteria justified by relevant development data, as applicable) are used to
1270 verify and validate the modification, (2) the results of verification and validation
1271 indicate that the performance is within the criteria, (3) the performance of the
1272 modified IVD has not significantly changed from the previously cleared performance
1273 claims, and (4) verification and validation do not reveal new risks or significantly
1274 modified existing risks apart from performance, then it is unlikely that the
1275 modification could significantly affect safety or effectiveness. If these criteria are
1276 met, then the modification is unlikely to significantly affect safety or effectiveness
1277 and manufacturers should proceed with the change making sure to document their
1278 assessment of whether a new 510(k) is required.

1279

Contains Nonbinding Recommendations

Draft – Not for Implementation

1280 If the results of routine verification and validation produce any unexpected issues or
1281 otherwise prove inadequate to verify and/or validate the modification—for example,
1282 pre-specified criteria are not met or the device fails to perform as expected—it is
1283 likely that the change could significantly affect the IVD’s safety and effectiveness,
1284 and a new 510(k) is likely required. This might be the case, for example, if the
1285 change necessitates a different verification and/or validation scheme.

1286
1287 Further, if non-standard verification or validation test methods or new acceptance
1288 criteria are necessary to produce the expected results, it is likely that the change could
1289 significantly affect safety or effectiveness and that a new 510(k) is required.

1290
1291 Even when the performance of the modified IVD falls within previously cleared
1292 performance claims, if the modified IVD’s performance specifications deviate from
1293 the performance values of widely accepted voluntary standards, that information
1294 should always be communicated to potential users in the labeling.

1295

1296 **E. Considerations for Risk Assessments of Modified**
1297 **Devices**

1298

1299 As discussed throughout this document, a device modification that leads to a significant
1300 change in the device’s risk profile likely requires a new 510(k). This section provides
1301 guidance on the principal factors to consider in conducting a risk assessment to determine
1302 whether a device modification leads to a significant change in the device’s risk profile.
1303 Manufacturers should use the risk assessment considerations discussed below in conjunction
1304 with the logic schemes and decision-making flowcharts outlined above.

1305

1306 FDA recommends that manufacturers use an accepted method of risk assessment, such as
1307 ISO 14971, an FDA-recognized standard that provides a framework for systematically
1308 managing risks of medical devices throughout the total product life cycle.

1309

1310 In general, the assessment of risk in deciding whether to submit a new 510(k) should identify
1311 all possible risks, and then focus on risks whose existence and characteristics are supported
1312 by objective scientific evidence. It is not necessary to focus on hypothetical risks that are not
1313 supported by scientific evidence or those that are determined to be negligible due to both the
1314 low probability of occurrence and low severity of harm. The manufacturer should then
1315 explore the severity and probability of occurrence of the harm to determine whether the
1316 device modification could significantly affect safety or effectiveness and require a new
1317 510(k).

1318

1319 Relationship between hazards and harm

1320

1321 Risk assessment involves describing the relationships between a hazard (a potential source of
1322 harm) and the ultimate consequences in terms of physical injury or damage. As part of their

Contains Nonbinding Recommendations

Draft – Not for Implementation

1323 risk assessment, manufacturers should analyze possible sequences of events, hazardous
1324 situations, and associated possible harm. This may include:

1325

- 1326 • initiating hazards, failure modes, or circumstances;
- 1327 • the sequences of events that could lead to a hazardous situation occurring;
- 1328 • the likelihood of such situations arising;
- 1329 • the likelihood that the hazardous situations lead to harm; and
- 1330 • the nature of the harm that could result.

1331

1332 The extent of risks and harms associated with a device modification may be assessed by
1333 taking into account the following factors, individually and in aggregate:

1334

1335 **1. Likelihood or probability of harm**

1336

1337 Various approaches may be employed to estimate probabilities of hazardous situations in
1338 assessing risk, including, but not limited to:

1339

- 1340 • use of relevant historical data;
- 1341 • prediction of probabilities of risk using analytical or simulation techniques;
- 1342 • reliability estimates;
- 1343 • production data; or
- 1344 • use of expert judgment.

1345

1346 The use of multiple approaches may be considered as this might serve to increase confidence
1347 in the results. Where uncertainty exists around these estimates, it may be useful to consider a
1348 qualitative approach to risk probability analysis. See, for instance, Section D.3 Risk
1349 Estimation of ISO 14971:2007 (second edition).

1350

1351 If it's determined that the likelihood of a harm occurring due to a device modification is
1352 negligible, then that change is unlikely to require a new 510(k). If it cannot be determined
1353 that a harm's likelihood is negligible, or the probability cannot be determined at all, then the
1354 below factors should also be considered.

1355

1356 **2. Severity of harm**

1357

1358 Manufacturers should consider the following points in analyzing the severity of a potential
1359 harm (refer to ISO 14971:2007 (second edition), Annex D, Sections D.3.3 and D.4 on
1360 severity and risk acceptability):

1361

- 1362 • New risks – If a device modification creates a new risk – i.e., a new hazard or
1363 hazardous situation – that did not exist for the original device and the new risk cannot
1364 be determined to be negligible, it is likely that the modification could significantly
1365 affect the device's safety or effectiveness, and a new 510(k) is likely required. An
1366 exception is a device change where the pre-mitigation risk level (the risk level before

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1367 any risk mitigations or controls are accounted for or product specifications are set)
1368 associated with the new risk is considered to be acceptable.
- 1369 • Changes in risk acceptability – If a device modification positively or negatively
1370 changes the pre-defined acceptability level (e.g., tolerable, acceptable, insignificant)
1371 of an individual risk based on the risk assessment, either before or after risk
1372 mitigation or control, it is likely that the modification could significantly affect the
1373 device’s safety or effectiveness, and a new 510(k) is likely required.
 - 1374 • Changes in risk score – In cases where there is no risk acceptability change for an
1375 affected risk, a major change to the severity score may still suggest potential
1376 significant impact to safety, depending on how the manufacturer determines their risk
1377 scores and defines risk acceptability. These types of changes will be very dependent
1378 on how a manufacturer conducts risk management and defines risk scores and risk
1379 acceptability.
 - 1380 • Duration – Some device features expose patients and/or users to temporary, minor
1381 harm; some can cause repeated but reversible harm; others can cause permanent,
1382 debilitating injury. Duration – that is, how long the adverse consequence lasts –
1383 should be considered along with the other factors described in this section.

1384
1385 Note that if a device change results in risk that could significantly affect the safety or
1386 effectiveness of a device, a new 510(k) must be submitted per 21 CFR 807.81(a)(3)(i), even
1387 if the risk can be mitigated.

1388 1389 **3. Device effectiveness**

1390
1391 Although ISO 14971 defines risk in terms of device harms and their effects on safety, it is
1392 important to note that whether a new 510(k) is required depends on whether the change could
1393 significantly affect the safety *or effectiveness* of the device. Therefore, manufacturers should
1394 also consider the possible effects a device modification may have on device effectiveness. As
1395 with safety risks, the manufacturer should consider the probability and severity (i.e.,
1396 magnitude) of impacts to device effectiveness.

1397
1398 In considering a device modification’s effects on device effectiveness, manufacturers should
1399 understand the criticality of the device feature being modified to the safe and effective use of
1400 the device. Certain features are more critical than others. For instance, the outer case of a
1401 ventilator, although important to the overall design of the device and providing for
1402 connection of various parts, is not as critical to the safe and effective use of the ventilator as
1403 the pump that circulates air to the patient. Note that labeling, which affects user actions, can
1404 be critical as well.

1405
1406

Appendix A: Examples

1407

1408

1409 The following are hypothetical examples of device changes with explanations as to why they
1410 likely would or would not require a new 510(k). These examples are intended to be
1411 illustrative of the thought process for different types of changes. Note that these generalized
1412 examples do not necessarily account for every possible detail, risk, or consideration a
1413 manufacturer should evaluate, and should not be taken to mean that the changes described
1414 definitely do or do not require a new 510(k). Real-world device modification decisions will
1415 depend on the particular details of the change and the specific device in question. Also note
1416 that devices with changes requiring a new 510(k) may not be legally commercially
1417 distributed before FDA clears the changed device. See 21 CFR 807.100(a) and sections
1418 513(f)(1) and 513(i) of the FD&C Act.

1419

1420

Labeling change examples

1421

1422 **1. Change:** The manufacturer of an IVD updates their labeling by changing the device from
1423 prescription use only to over-the-counter use.

1424

Relevant questions:

1425

A1– *Is it a substantive change in the indications for use?* Yes. The revised labeling
1426 expands the scope of intended users of the device to untrained users, which typically
1427 could significantly affect the safety or effectiveness of the device.

1428

Decision: Submit the change in a new 510(k).

1429

1430 **2. Change:** The manufacturer of a device adds a precaution stating that the device must be
1431 properly sterilized prior to use for patient safety. The modified labeling does not modify
1432 the cleaning, disinfection, or sterilization instructions.

1433

Relevant questions:

1434

A3 – *Is it a change in warnings or precautions?* Yes. Proceed to A5.1.

1435

A5.1 – *Could the change affect the indications for use?* No. The added precaution simply
1436 emphasizes proper sterilization and does not affect the indications for use.

1437

A5.2 – *Does a risk assessment of the changed device identify any new risks or
1438 significantly modified existing risks?* No. The added precaution simply emphasizes
1439 proper sterilization and does not affect the device’s risk profile.

1440

Decision: Document the change to file.

1441

1442

3.

1443

a. Change: The manufacturer of an IVD removes a limitation contained in
1444 their labeling that informs users that heterophilic human anti-mouse antibodies
1445 (HAMA) cause interference in their assay, which can lead to false results that could
1446 harm the end-user. The manufacturer removes this limitation without making any
1447 changes to the assay itself.

1448

Relevant questions:

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1449 A3– *Is it a change in warnings or precautions?* Yes. This change removed the
1450 statement from the limitation section of the labeling that HAMA may cross-react with
1451 the assay. Proceed to A5.1.
1452 A5.1– *Could the change affect the indications for use?* No. The limitation warns users
1453 about potential cross-reactivity and does not affect the indications for use.
1454 A5.2 – *Does a risk assessment of the changed device identify any new risks or*
1455 *significantly modified existing risks?* Yes. Removing an identified interference from
1456 the labeling could lead to falsely elevated or falsely low analyte concentration,
1457 depending on the site of the interference in the immunoassay reaction. The removal of
1458 the limitation may result in the user failing to be alerted to a known risk and may
1459 impact performance by changing the ability to accurately measure the analyte
1460 concentration.
1461 **Decision:** Submit the change in a new 510(k).
1462
1463 **b. Change:** The manufacturer of an IVD updates their labeling by adding a new
1464 limitation after identifying a newly approved drug as a potential interferent.
1465 **Relevant questions:**
1466 Main flowchart, question 1 – *Change made with intent to significantly improve the*
1467 *safety or effectiveness of the device, e.g., in response to a known risk, adverse event,*
1468 *etc.?* No. The manufacturer is only aware that the newly approved drug may cause
1469 interference with their assay and has not received any reports of adverse events. The
1470 labeling change is made to add the new limitation.
1471 A3– *Is it a change in warnings or precautions?* Yes. The change adds a new limitation
1472 to the IVD labeling and the manufacturer has monitored device usage and updated the
1473 labeling accordingly. Proceed to A5.1.
1474 A5.1– *Could the change affect the indications for use?* No. The interferent does not
1475 affect the indications for use for this particular device.
1476 A5.2 – *Does a risk assessment of the changed device identify any new risks or*
1477 *significantly modified existing risks?* No. The labeling change does not significantly
1478 affect the device’s risk profile because no new risks or significantly modified existing
1479 risks are identified in the risk assessment.
1480 **Decision:** Document the change to file.
1481
1482 **4. Change:** The warning information in the labeling for an IVD is modified to account for
1483 recently revised hazardous material guidelines.
1484 **Relevant questions:**
1485 A3 – *Is it a change in warnings or precautions?* Yes. A change is made to a warning
1486 about hazardous materials. Proceed to A5.1
1487 A5.1 – *Does the change affect the indications for use?* No. The updated warning
1488 information does not affect the device’s indications for use.
1489 A5.2 – *Does a risk assessment of the changed device identify any new risks or*
1490 *significantly modified existing risks?* No. So long as the same risks are communicated to
1491 the device user, this change would not significantly affect the device’s risk profile.
1492 **Decision:** Document the change to file.
1493

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1494 **5. Change:** The manufacturer adds a foreign language translation of the instructions for use
1495 to a device’s labeling. The translation does not change the meaning of the instructions.
1496 **Relevant questions:**
1497 *A4 – Does the change affect the instructions for use or other pieces of the labeling? Yes.*
1498 *A5.1 – Could the change affect the indications for use? No; as long as the translation*
1499 *does not change the meaning of the instructions, this change would not affect the*
1500 *indications for use.*
1501 *A5.2 – Does a risk assessment of the changed device identify any new risks or*
1502 *significantly modified existing risks? No. Again, as long as the translation does not*
1503 *change the meaning of the instructions, this change would not affect the device’s risk*
1504 *profile.*
1505 **Decision:** Document the change to file.
1506
- 1507 **6. Change:** The instructions for use of a catheter guidewire are modified to provide
1508 instructions on how to access different types of vasculature that were not previously
1509 addressed in the labeling.
1510 **Relevant questions:**
1511 *A4 – Does the change affect the instructions for use or other pieces of the labeling? Yes.*
1512 *A5.1 – Could the change affect the indications for use? Yes. The revised instructions*
1513 *suggest that the device can be used in new vasculature, which would be considered an*
1514 *expansion of the device’s indications for use, which could significantly affect safety and*
1515 *effectiveness.*
1516 **Decision:** Submit the change in a new 510(k).
1517
- 1518 **7.**
- 1519 **a. Change:** The original instructions for use for a surgical laser intended to treat stones
1520 in the urinary tract only included instructions on lithotripsy modes. The instructions
1521 are modified to provide instructions on ablating soft tissue.
1522 **Relevant questions:**
1523 *A4 – Does the change affect the instructions for use or other pieces of the labeling?*
1524 *Yes.*
1525 *A5.1 – Could the change affect the indications for use? Yes. The revised instructions*
1526 *could result in the device being used for ablation of soft tissue, which would be a new*
1527 *indication for use that could result in new device risks.*
1528 **Decision:** Submit the change in a new 510(k).
1529
- 1530 **b. Change:** The original instructions for use for a surgical laser intended to treat stones
1531 in the urinary tract only included instructions on lithotripsy modes. The instructions
1532 are modified to provide additional instructions on the existing settings for lithotripsy
1533 on the cleared device, and does not modify instructions regarding compatible
1534 procedures or instruments.
1535 **Relevant questions:**
1536 *A4 – Does the change affect the instructions for use or other pieces of the labeling?*
1537 *Yes.*

Contains Nonbinding Recommendations

Draft – Not for Implementation

1538 A5.1 – *Could the change affect the indications for use?* No. This change does not
1539 affect the indications for use. The device was cleared with indications for lithotripsy;
1540 the change only clarifies the settings.
1541 A5.2 – *Does a risk assessment of the changed device identify any new risks or*
1542 *significantly modified existing risks?* No. The manufacturer’s risk assessment
1543 concludes that the clarification of already existing settings does not introduce any new
1544 device risks, and the risk acceptability for the previously existing risks is not changed.
1545 **Decision:** Document the change to file.
1546

1547 **8. Change:** A manufacturer changes the design of an IVD for diagnosing herpes simplex 1
1548 and 2 to a less strict performance specification that decreases both the sensitivity and
1549 specificity of the device to increase production. The manufacturer updates the
1550 performance specifications found in the labeling of the device.

1551 **Relevant questions:**

1552 A4 – *Does the change affect the instructions for use or other pieces of the labeling?* Yes.
1553 A5.1 – *Could the change affect the indications for use?* No. The device is still indicated
1554 for the same use.

1555 A5.2- *Does a risk assessment of the changed device identify any new risks or*
1556 *significantly modified existing risks?* Yes. The modifications to the device result in
1557 significantly increased existing risks. This is due to a mathematically expected increase in
1558 false positive results, which would, in turn, be expected to lead to an increase in harms
1559 such as mental anguish, delayed diagnosis for the true cause of any symptoms, and
1560 unnecessary treatment (e.g., pregnant women and newborns receiving unnecessary
1561 antiviral drugs or an unnecessary caesarean delivery of the fetus). Further, this would
1562 also significantly increase risks due to a mathematically expected increase in false
1563 negative results, which would, in turn, be expected to lead to an increase in harms such as
1564 delayed diagnosis that would in turn delay treatment of the underlying condition and
1565 could lead to unintended spread of the disease (e.g., through sexual partners, neonatal
1566 transmission during vaginal delivery, and transplanted organs).

1567 **Decision:** Submit the change in a new 510(k).

1568 **Note:** This type of change in labeling is in response to a design change. Accordingly,
1569 analyses under Section A and Section D would apply. See Example 34.
1570

1571

1572

Design change examples

1573

1574 **9. Change:** A device is modified to use an internal battery instead of an external AC power
1575 source.

1576 **Relevant questions:**

1577 B2 – *Is it a control mechanism, operating principle, or energy type change?* Yes. This is
1578 an energy type change, which typically requires a new 510(k) due to the likelihood of
1579 such a change to significantly affect safety or effectiveness.

1580 **Decision:** Submit the change in a new 510(k).
1581

Contains Nonbinding Recommendations

Draft – Not for Implementation

1582 **10. Change:** The manufacturer changes the packaging for their device, which is provided
1583 sterile, from one variant of polyethylene to another due to a material supplier change. An
1584 analysis shows the new polyethylene has no impurities that could affect the device's
1585 biocompatibility. The manufacturer will use the same package integrity test protocol as
1586 the one described in its previously cleared 510(k) to support the change.

1587 **Relevant questions:**

1588 B4 – *Is there a change in packaging or expiration dating?* Yes.

1589 B4.1 – *Is the same method or protocol, as described in a previously cleared 510(k) used*
1590 *to support the change?* Yes.

1591 **Decision:** Document the change to file.

1592

1593 **11.**

1594 **a. Change:** A biliary stent manufacturer adds a new larger stent diameter to a family of
1595 biliary stents, 1 mm outside of the range of the manufacturer's previously cleared
1596 stents. The stent lengths are unchanged.

1597 **Relevant questions:**

1598 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1599 *wireless communication, components or accessories, or the patient/user interface)?*
1600 Yes.

1601 B5.1 – *Does the change significantly affect the use of the device?* The answer to this
1602 question depends on the original diameter of the stent and the extent of change in the
1603 diameter.

1604 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1605 *significantly modified existing risks?* Yes. The diameter of a biliary stent is critical to
1606 the device's safety and effectiveness. The greater stent diameter significantly affects
1607 device-related safety and effectiveness risks.

1608 **Decision:** Submit the change in a new 510(k).

1609

1610 **b. Change:** A biliary stent manufacturer adds a new stent diameter to a family of stents,
1611 within the range of the diameters of the manufacturer's previously cleared stents. The
1612 stent lengths are unchanged. The previously cleared 510(k) for the stents objectively
1613 demonstrated that the smallest and largest stent diameters (the minimum and
1614 maximum ends of the diameter size range) were the worst-case scenarios in terms of
1615 the safety and effectiveness risks.

1616 **Relevant questions:**

1617 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1618 *wireless communication, components or accessories, or the patient/user interface)?*
1619 Yes.

1620 B5.1 – *Does the change significantly affect the use of the device?* No. Because the new
1621 diameter is within the range of the previously cleared stents, the manufacturer
1622 determines that the change does not significantly affect the use of the device.

1623 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1624 *significantly modified existing risks?* No. Since the new stent diameter is within the
1625 range of the manufacturer's previously cleared stents of the same lengths, and the
1626 previously cleared 510(k) objectively demonstrated that the smallest and largest

Contains Nonbinding Recommendations

Draft – Not for Implementation

1627 diameter sizes represented worst-case scenarios in terms of the safety and
1628 effectiveness risks, the new diameter would not significantly affect the risk profile of
1629 the device.

1630 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
1631 *design validation?* No. The manufacturer determines clinical data are not necessary for
1632 their specific change. They make the initial decision at this point to document the
1633 change to file.

1634 B5.4 – *Do design verification and/or validation activities produce any unexpected*
1635 *issues of safety or effectiveness?* No. In this example, routine verification and
1636 validation activities are conducted successfully.

1637 **Decision:** Document the change to file.
1638

1639 **12. Change:** In order to better accommodate connection of a urinary drainage (Foley)
1640 catheter to a collection apparatus, the length of the catheter is increased by several
1641 millimeters. The new length is outside of previously cleared lengths for this device.

1642 **Relevant questions:**

1643 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1644 *wireless communication, components or accessories, or the patient/user interface)?* Yes.

1645 B5.1 – *Does the change significantly affect the use of the device?* No. The device’s
1646 increased length would not suggest use of the device for purposes, locations, or
1647 populations other than those for which it was cleared, so the manufacturer determines that
1648 the change does not significantly affect the use of the device.

1649 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1650 *significantly modified existing risks?* Extreme length changes may affect the risk profile
1651 of a urinary drainage catheter (e.g., for biocompatibility), but in general, length changes
1652 for this device are unlikely to create new risks or significantly affect existing risks by
1653 affecting the acceptability of those risks. Device specifics will be important in this
1654 example, however, in this example the change does not significantly affect the device’s
1655 risk profile.

1656 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
1657 *design validation?* No. The manufacturer determines clinical data are not necessary for
1658 their specific change. They make the initial decision at this point to document the change
1659 to file.

1660 B5.4 – *Do design verification and/or validation activities produce any unexpected issues*
1661 *of safety or effectiveness?* No. In this example, routine verification and validation
1662 activities are conducted successfully.

1663 **Decision:** Document the change to file.
1664

1665 **13.**

1666 **a. Change:** The manufacturer of a urinary drainage (Foley) catheter reduces the diameter
1667 of the catheter to supplement a family of catheters. The new diameter is within the
1668 range of previously cleared diameters for this device, and the previously cleared
1669 510(k) objectively demonstrated the smallest and largest diameters to be worst-case
1670 scenarios in terms of the safety and effectiveness risks. The new diameter is within the
1671 range of sizes used for smaller adult patients for increased comfort.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1672
1673
1674
1675
1676
1677
1678
1679
1680
1681
1682
1683
1684
1685
1686
1687
1688
1689
1690
1691
1692
1693
1694
1695
1696
1697
1698
1699
1700
1701
1702
1703
1704
1705
1706
1707
1708
1709
1710
1711
1712
1713
1714
1715
1716

Relevant questions:

B5 – *Is it any other change in design (e.g., dimensions, performance specifications, wireless communication, components or accessories, or the patient/user interface)?* Yes.

B5.1 – *Does the change significantly affect the use of the device?* No. This new catheter size would be expected to be used in the same patient population as the previously cleared devices.

B5.2 – *Does a risk assessment of the changed device identify any new risks or significantly modified existing risks?* No. Since the modified device is within the currently cleared range of dimensions and the smallest and largest previously cleared sizes were demonstrated to be worst-case scenarios in terms of the safety and effectiveness risks, this change would not significantly affect the risk profile of the device.

B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of design validation?* No. The manufacturer determines clinical data are not necessary for their specific change. They make the initial decision at this point to document the change to file.

B5.4 – *Do design verification and/or validation activities produce any unexpected issues of safety or effectiveness?* No. In this example, routine verification and validation activities are conducted successfully.

Decision: Document the change to file.

- b. Change:** The manufacturer of a urinary drainage (Foley) catheter reduces the diameter of the catheter. The new diameter is outside of the range of previously cleared diameters for this device. The new diameter is also smaller than what is typically used for adult patients, and is of a size that is typically used for pediatric patients. The device is not cleared for pediatric use.

Relevant questions:

B5 – *Is it any other change in design (e.g., dimensions, performance specifications, wireless communication, components or accessories, or the patient/user interface)?* Yes.

B5.1 – *Does the change significantly affect the use of the device?* Even if the indications for use and labeling are not changed, this new diameter significantly affects the use of the device by changing it from adult use to pediatric use. This could significantly affect the safety and effectiveness of the device.

Decision: Submit the change in a new 510(k).

- 14. Change:** The manufacturer of a biliary stent increases the thickness of the nitinol wire in the stent from that used in the previously cleared device to reduce potential for stent fractures.

Relevant questions:

B5 – *Is it any other change in design (e.g., dimensions, performance specifications, wireless communication, components or accessories, or the patient/user interface)?* Yes.

B5.1 – *Does the change significantly affect the use of the device?* No. The thickness of the nitinol wire of the device would not significantly affect its use.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1717 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1718 *significantly modified existing risks?* Yes. The thickness of the wire is critical to the
1719 performance of the stent, so an increase could significantly affect the risk profile and the
1720 safety or effectiveness of the device.

1721 **Decision:** Submit the change in a new 510(k).

1722

1723 **15. Change:** The manufacturer adds a foot switch to control an endoscopic electrosurgical
1724 unit. The previously cleared device did not have a foot switch.

1725 **Relevant questions:**

1726 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1727 *wireless communication, components or accessories, or the patient/user interface)?* Yes.
1728 This is a change to the device’s user interface.

1729 B5.1 – *Does the change significantly affect the use of the device?* No. The addition of a
1730 foot switch would not significantly affect the use of the device.

1731 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1732 *significantly modified existing risks?* Yes. The addition of the foot switch presents new
1733 risks; if it operates incorrectly it could cause the device to function incorrectly, which
1734 could significantly affect the safety and effectiveness of the device.

1735 **Decision:** Submit the change in a new 510(k).

1736

1737 **16. Change:** The grip portion of a diagnostic ultrasound transducer is redesigned to improve
1738 user comfort.

1739 **Relevant questions:**

1740 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1741 *wireless communication, components or accessories, or the patient/user interface)?* Yes.
1742 This is a change to the device’s user interface.

1743 B5.1 – *Does the change significantly affect the use of the device?* No. In this example, the
1744 redesign of the grip would not significantly affect the use of the device.

1745 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1746 *significantly modified existing risks?* No. Although the change to the transducer grip
1747 could affect certain risks, such as the user potentially mishandling the device, the severity
1748 of these risks for this device is low. (Note that mishandling a device such as a surgical
1749 instrument, however, would produce more severe risks, and could possibly lead to a new
1750 510(k) being required.)

1751 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
1752 *design validation?* No. The manufacturer determines clinical data are not necessary for
1753 their specific change. They make the initial decision at this point to document the change
1754 to file.

1755 B5.4 – *Do design verification and/or validation activities produce any unexpected issues*
1756 *of safety or effectiveness?* No. In this example, routine verification and validation
1757 activities are conducted successfully.

1758 **Decision:** Document the change to file.

1759

1760 **17. Change:** A particular device heats fluid in order to achieve its intended effect. The most
1761 recently cleared device had a low-power heater and the maximum fluid temperature was

Contains Nonbinding Recommendations

Draft – Not for Implementation

1762 low enough that the severity of the worst-case thermal injury was low to moderate. In the
1763 risk analysis for the design of the most recently cleared device, the risk score/rating for
1764 thermal injury was therefore in a range identified in the risk management document as
1765 “tolerable but undesirable,” before risk control measures were added. After receiving
1766 input from customers that the fluid heating process was too slow, the device was changed
1767 to use a higher-powered heater, which increased the maximum possible fluid
1768 temperature.

1769 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1770 *wireless communication, components or accessories, or the patient/user interface)?* Yes.

1771 B5.1 – *Does the change significantly affect the use of the device?* No. This change would
1772 not significantly affect the use of the device.

1773 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1774 *significantly modified existing risks?* Yes. When the manufacturer performed a risk
1775 analysis on the new design, the severity of potential thermal injury increased and the risk
1776 of thermal injury became “unacceptable,” before application of additional risk control
1777 measures. This risk analysis showed that the design change had a potentially significant
1778 impact on safety by changing the pre-mitigation acceptability of the risk. Therefore, a
1779 new 510(k) is likely required. This same conclusion holds whether or not the
1780 manufacturer needed to add new risk control measures to bring the final risk into the
1781 acceptable range.

1782 **Decision:** Submit the change in a new 510(k).
1783

1784 **18. Change:** A device includes a sharp edge in order to achieve the intended clinical effect.
1785 The manufacturer changed the device to include risk control measures to reduce the
1786 chance of unintended contact with the sharp edge, but those measures were only partially
1787 effective. In the original design, after the risk controls were in place, the risk score/rating
1788 for patient exposure to the sharp edge on the device was “tolerable but undesirable.” The
1789 manufacturer conducted a risk-benefit analysis that showed that the benefits of the device
1790 outweighed the risk associated with sharp edge exposure and therefore marketed the
1791 device. A subsequent design change was found to be more effective at preventing
1792 unintended sharp edge exposure. As a result, the risk score/rating was reduced and the
1793 post-mitigation risk was in the acceptable range.

1794 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1795 *wireless communication, components or accessories, or the patient/user interface)?* Yes.

1796 B5.1 – *Does the change significantly affect the use of the device?* No. This change would
1797 not significantly affect the use of the device.

1798 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1799 *significantly modified existing risks?* Yes. The proposed change changes the risk
1800 acceptability and severity, which yields a significant improvement in the device’s risk
1801 profile. The manufacturer concludes, therefore, that the change could significantly affect
1802 the safety or effectiveness of the device.

1803 **Decision:** Submit the change in a new 510(k).
1804

1805 **19. Change:** A device designed with moving parts has an inherent risk of pinching the user.
1806 The established risk control measure was a guard placed to prevent contact with the

Contains Nonbinding Recommendations

Draft – Not for Implementation

1807 pinching zone. The guard was highly effective. The occurrence rating for pinching after
1808 the risk control was added was “remote,” (defined by the manufacturer with a probability
1809 of occurrence of $<10^{-5}$ and $\geq 10^{-6}$). The manufacturer pre-defined that an acceptable risk
1810 analysis would determine that occurrence was “remote” or better. The manufacturer is
1811 considering changing the device in a way that would modify the dimensional tolerances
1812 of the guard for better manufacturability. Risk analysis related to the change in tolerances
1813 concluded that the severity of harm was unchanged. However, the probability of
1814 occurrence increased from “remote” to “improbable” (defined by the manufacturer with a
1815 probability of occurrence of $<10^{-6}$). The risk remained acceptable according to the
1816 predefined acceptability criteria in the risk management plan.

1817 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1818 *wireless communication, components or accessories, or the patient/user interface)?* Yes.

1819 B5.1 – *Does the change significantly affect the use of the device?* No. This change would
1820 not significantly affect the use of the device.

1821 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1822 *significantly modified existing risks?* No. In this case, there are no new risks, and even
1823 though the probability of the risk in question increases slightly, the overall pre-defined
1824 acceptability category of the risk is unchanged, so the changes to the risk profile are not
1825 significant.

1826 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
1827 *design validation?* No. The manufacturer determines clinical data are not necessary for
1828 their specific change. They make the initial decision at this point to document the change
1829 to file.

1830 B5.4 – *Do design verification and/or validation activities produce any unexpected issues*
1831 *of safety or effectiveness?* No. In this example, routine verification and validation
1832 activities are conducted successfully.

1833 **Decision:** Document the change to file.

1834

1835 **20.**

1836 **a. Change:** A portable medical device receives its power through a removable,
1837 rechargeable battery. The device manufacturer provides a battery charging station for
1838 the battery. The proposed change is to the design of the battery charging station. There
1839 is no change in the battery itself, only the means by which it is charged. The device is
1840 not life-sustaining or life-supporting.

1841 **Relevant questions:**

1842 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1843 *wireless communication, components or accessories, or the patient/user interface)?*
1844 Yes.

1845 B5.1 – *Does the change significantly affect the use of the device?* No. This change
1846 would not significantly affect the use of the device.

1847 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1848 *significantly modified existing risks?* No. Because the device can operate without the
1849 battery charging station, the battery itself is easily replaced, and the device is not life-
1850 sustaining or life-supporting, the severities of risks surrounding the battery charging
1851 station are low. Unless any new risks are associated with the change or the likelihood

Contains Nonbinding Recommendations

Draft – Not for Implementation

1852 of risks associated with the battery charging station are significantly increased, this
1853 change would not significantly affect the device’s risk profile.
1854 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
1855 *design validation?* No. The manufacturer determines clinical data are not necessary for
1856 their specific change. They make the initial decision at this point to document the
1857 change to file.
1858 B5.4 – *Do design verification and/or validation activities produce any unexpected*
1859 *issues of safety or effectiveness?* No. In this example, routine verification and
1860 validation activities are conducted successfully.

1861 **Decision:** Document the change to file.
1862

1863 **21. Change:** A manufacturer changes the surface of a titanium dental implant from an
1864 untreated surface to one that is acid-etched. The surface is in direct contact with the
1865 patient’s bone. The manufacturer has not previously used the acid-etching process, and a
1866 cleaning process is necessary to remove acid from the device surface.

Relevant questions:

1868 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
1869 *wireless communication, components or accessories, or the patient/user interface)?* Yes.
1870 This a design change because the implant’s surface properties are changed.

1871 B5.1 – *Does the change significantly affect the use of the device?* No. This change would
1872 not significantly affect the use of the device.

1873 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
1874 *significantly modified existing risks?* Yes. Surface changes can significantly affect the
1875 safety and effectiveness of an implant by, for example, significantly modifying the
1876 likelihood of implant instability. This can be considered a safety risk, and since the
1877 interaction between the implant and the *in vivo* environment is critical to the stability of
1878 the implant and therefore its effectiveness, this could also be considered a significant
1879 impact on the device’s effectiveness.

1880 **Decision:** Submit the change in a new 510(k).

1881 **Note:** This change could also be evaluated as a materials change. See Example 27.
1882

Materials change examples

1883
1884
1885 **22. Change:** The manufacturer of a catheter changes its supplier that provides the polymer
1886 tubing used to manufacture the catheter. The manufacturer conducts chemical
1887 characterization tests that show the new supplier’s polymer is nearly identical to the
1888 original supplier’s. The assessment shows there are no new components, and that the
1889 additional amounts of some components are not likely to affect the biocompatibility of
1890 the finished device.

Relevant questions:

1892 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1893 *the material’s processing?* Yes. The material provided by the new supplier is slightly
1894 different than that provided by the original supplier.

1895 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?* Yes.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1896 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?* No.
1897 The manufacturer has conducted a risk assessment that demonstrates the changes in
1898 material formulation between the original supplier’s and the new supplier’s polymers are
1899 minor and will not affect the biocompatibility of the finished device.

1900 C5 – *Could the change affect the device’s performance specifications?* No. For the
1901 purposes of this example, the manufacturer’s assessment shows that the differences in
1902 formulation are minor and not likely to affect the performance of the finished device.

1903 **Decision:** Document the change to file.

1904

1905 **23.**

1906 **a. Change:** The manufacturer of a catheter changes the material of its catheter from
1907 polymer A to polymer B. The manufacturer has not previously used polymer B in any
1908 of its devices, but knows of another catheter on the market with the same cleared
1909 indications for use that uses polymer B.

1910 **Relevant questions:**

1911 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1912 *the material’s processing?* Yes.

1913 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
1914 Yes.

1915 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*

1916 Yes. Polymer B has a different chemical formulation that may affect the
1917 biocompatibility of the catheter.

1918 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
1919 *device?* No. Even though there is another catheter on the market made of polymer B,
1920 the other device may have a different formulation or different manufacturing or
1921 finishing processes that could affect the biocompatibility or performance.

1922 **Decision:** Submit the change in a new 510(k).

1923

1924 **b. Change:** The manufacturer of a catheter changes the material of its catheter from
1925 polymer A to polymer B. The manufacturer has used the same polymer B, with the
1926 same formulation and processing, in another cleared model of catheter with the same
1927 type and duration of contact and the same performance specifications.

1928 **Relevant questions:**

1929 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1930 *the material’s processing?* Yes.

1931 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
1932 Yes.

1933 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*

1934 Yes. Polymer B has a different chemical formulation that may affect the
1935 biocompatibility of the catheter.

1936 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
1937 *device?* Yes. The manufacturer has used the same polymer B, with the same
1938 formulation and processing, in another model of catheter with the same type and
1939 duration of contact. This addresses the possible biocompatibility concerns.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1940 C5 – *Could the change affect the device’s performance specifications?* No. The
1941 manufacturer has used the same polymer B in another model of catheter with the same
1942 performance specifications.
1943 **Decision:** Document the change to file.
1944
- 1945 **c. Change:** A manufacturer changes the material of its catheter, intended for prolonged
1946 blood contact, from polymer A to polymer B. The manufacturer has used the same
1947 polymer B in another cleared device; however, this other device was indicated for a
1948 use with limited duration and skin contact only.
1949 **Relevant questions:**
1950 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1951 *the material’s processing?* Yes.
1952 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
1953 Yes.
1954 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*
1955 Yes. Polymer B has a different chemical formulation that may affect the
1956 biocompatibility of the catheter.
1957 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
1958 *device?* No. The manufacturer has used the same polymer B, with the same
1959 formulation and processing, in another device, however, the other device was subject
1960 to a less risky type and duration of contact. The modified device will be subjected to
1961 additional biocompatibility risks compared to the other polymer B device, and
1962 therefore the use of polymer B in the other device does not address the
1963 biocompatibility concerns.
1964 **Decision:** Submit the change in a new 510(k).
1965
- 1966 **d. Change:** A manufacturer changes the material of a device intended for limited skin
1967 contact from polymer A to polymer B. The manufacturer has used the same polymer B
1968 in another cleared device that was intended for prolonged blood contact and had the
1969 same performance specifications.
1970 **Relevant questions:**
1971 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1972 *the material’s processing?* Yes.
1973 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
1974 Yes.
1975 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*
1976 Yes. Polymer B has a different chemical formulation that may affect the
1977 biocompatibility of the catheter.
1978 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
1979 *device?* Yes. The manufacturer has used the same polymer B, with the same
1980 formulation and processing, in another cleared device with a riskier type and duration
1981 of contact, and the size and geometry of the new device would not affect curing of the
1982 polymer or result in more material in the new device. The riskier use of the material in
1983 the other cleared device shows that the polymer B can be expected to be biocompatible
1984 in its new application.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1985 C5 – *Could the change affect the device’s performance specifications?* No. The
1986 manufacturer used the same polymer B in another model of catheter with the same
1987 performance specifications.
1988 **Decision:** Document the change to file.
1989
- 1990 **24. Change:** A manufacturer changes the material of a catheter from material A to material
1991 B, which is used in another of the manufacturer’s cleared catheters. Material A is molded,
1992 and material B, used in the other catheter, is extruded.
1993 **Relevant questions:**
1994 C2 – *Is this a change in material type, material formulation, chemical composition, or*
1995 *the material’s processing?* Yes.
1996 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?* Yes.
1997 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?* Yes.
1998 The new material B has a different chemical formulation than the original material A that
1999 may affect the biocompatibility of the device.
2000 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
2001 *device?* No. The manufacturer has used the same material in another cleared catheter, but
2002 the processing of the material is different, which may affect biocompatibility.
2003 **Decision:** Submit the change in a new 510(k).
2004
- 2005 **25.**
2006 **a. Change:** A manufacturer decides to change the material of a catheter from material A
2007 to material B. Material B is used in another of the manufacturer’s own cleared
2008 catheters with similar type and duration of patient contact. Material A is sterilized by
2009 gamma irradiation, and material B is sterilized by ethylene oxide.
2010 **Relevant questions:**
2011 C2 – *Is this a change in material type, material formulation, chemical composition, or*
2012 *the material’s processing?* Yes.
2013 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
2014 Yes.
2015 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*
2016 Yes. Material B has a different chemical formulation than material A that may affect
2017 the biocompatibility of the device.
2018 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
2019 *device?* No. The manufacturer has used material B in another cleared catheter, but the
2020 processing of the material is different, which may affect biocompatibility.
2021 **Decision:** Submit the change in a new 510(k).
2022
- 2023 **b. Change:** A manufacturer decides to change the material of a catheter from material A
2024 to material B. Material B is used in another of the manufacturer’s own cleared
2025 catheters, which has the same type and duration of patient contact, as well as the same
2026 performance specifications. Both materials A and B are molded and are sterilized by
2027 ethylene oxide.
2028 **Relevant questions:**

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 2029 C2 – *Is this a change in material type, material formulation, chemical composition, or*
2030 *the material’s processing?* Yes.
- 2031 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
2032 Yes.
- 2033 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*
2034 Yes. Material B has a different chemical formulation than material A that may affect
2035 the biocompatibility of the device.
- 2036 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
2037 *device?* Yes. The manufacturer has used material B in another cleared catheter, and the
2038 processing is the same. In addition, the size and geometry of the new device would not
2039 affect curing of the polymer or result in more material in the new device, and there are
2040 no differences in how material B is joined to other components of the catheter (e.g.,
2041 type of adhesive, or conditions of heat welding) that could result in different
2042 interactive chemistries.
- 2043 C5 – *Could the change affect the device’s performance specifications?* No. The
2044 manufacturer has used the same material B in another model of catheter with the same
2045 performance specifications, which is processed in the same manner.
- 2046 **Decision:** Document the change to file.
- 2047
- 2048 **c. Change:** A manufacturer decides to change the material of a catheter from material A
2049 to material B. Material B is used in another of the manufacturer’s own cleared
2050 catheters, which has the same type and duration of patient contact, but different
2051 performance specifications. Both materials A and B are molded and are sterilized by
2052 ethylene oxide.
- 2053 **Relevant questions:**
- 2054 C2 – *Is this a change in material type, material formulation, chemical composition, or*
2055 *the material’s processing?* Yes.
- 2056 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?*
2057 Yes.
- 2058 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?*
2059 Yes. Material B has a different chemical formulation than material A that may affect
2060 the biocompatibility of the device.
- 2061 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
2062 *device?* Yes. The manufacturer has used material B in another cleared catheter, and the
2063 processing is the same. In addition, the size and geometry of the new device would not
2064 affect curing of the polymer or result in more material in the new device, and there are
2065 no differences in how material B is joined to other components of the catheter (e.g.,
2066 type of adhesive, or conditions of heat welding) that could result in different
2067 interactive chemistries.
- 2068 C5 – *Could the change affect the device’s performance specifications?* Yes. The
2069 manufacturer used the same material B in another model of catheter; however, the
2070 performance specifications were different. The new material could potentially affect
2071 the device’s performance, so the manufacturer is directed to B4.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 2072 B5 – *Is it any other change in design (e.g., dimensions, performance specifications,*
2073 *wireless communication, components or accessories, or the patient/user interface,)?*
2074 Yes.
- 2075 B5.1 – *Does the change significantly affect the use of the device?* No. The new
2076 material does not significantly affect the use of this device.
- 2077 B5.2 – *Does a risk assessment of the changed device identify any new risks or*
2078 *significantly modified existing risks?*
- 2079 If the new material has significantly different physical properties than the material in
2080 the previously cleared device, the risk profile of the device could be significantly
2081 affected in terms of risk score, risk acceptability, etc., and a new 510(k) may be
2082 required. However, for the purposes of this example, the new material is not expected
2083 to have significantly different physical properties, so a 510(k) would not be required.
- 2084 B5.3 – *Are clinical data necessary to evaluate safety or effectiveness for purposes of*
2085 *design validation?* No. The manufacturer determines clinical data are not necessary for
2086 their specific change. They make the initial decision at this point to document the
2087 change to file.
- 2088 B5.4 – *Do design verification and/or validation activities produce any unexpected*
2089 *issues of safety or effectiveness?* No. In this example, routine verification and
2090 validation activities are conducted successfully.
- 2091 **Decision:** Document the change to file.
2092
- 2093 **26. Change:** The manufacturer of a dental implant changes the surface of a titanium dental
2094 implant from an untreated surface to one that is acid-etched. The surface is in direct
2095 contact with the patient’s bone. The manufacturer has not previously used the acid-
2096 etching process, and a cleaning process is necessary to remove acid from the device
2097 surface.
- 2098 **Relevant questions:**
- 2099 C2 – *Is this a change in material type, material formulation, chemical composition, or*
2100 *the material’s processing?* Yes. The material processing of the device has been changed.
- 2101 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?* Yes.
- 2102 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?* Yes.
2103 Residue from the acid-etching process may affect the biocompatibility of the device.
- 2104 C4.1 – *Has the manufacturer used the same material in a similar legally marketed*
2105 *device?* No. The manufacturer has not previously used the acid-etching process.
- 2106 **Decision:** Submit the change in a new 510(k).
2107 **Note:** This change could also be evaluated as a design change. See Example 22.
2108
- 2109 **27. Change:** The manufacturer of an implantable device applies a temporary tape to the
2110 device for identification of manufacturing steps. The tape has been demonstrated in peer-
2111 reviewed literature to not leave adhesive on the surface of the device.
- 2112 **Relevant questions:**
- 2113 C2 – *Is this a change in material type, material formulation, chemical composition, or*
2114 *the material’s processing?* Yes. The material processing of the device has been changed.
- 2115 C3 – *Will the changed material directly or indirectly contact body tissues or fluids?* Yes.

Contains Nonbinding Recommendations

Draft – Not for Implementation

2116 C4 – *Does a risk assessment identify any new or increased biocompatibility risks?* No.
2117 The tape has been demonstrated to not leave adhesive on the surface of the device.
2118 C5 – *Could the change affect the device’s performance specifications?* No. The tape is
2119 temporary for manufacturing purposes, and is removed before clinical use of the device.
2120 Since the tape has been demonstrated to not leave adhesive on the surface of the device, it
2121 would not be expected to affect the device’s performance.
2122 **Decision:** Document the change to file.
2123

IVD technology, engineering, performance, and materials change examples

2124

2125

2126

2127

2128

2129

2130

2131

2132

2133

2134

2135

2136

2137

2138

2139

2140

2141

2142

2143

2144

2145

2146

2147

2148

2149

2150

2151

2152

2153

2154

2155

2156

2157

2158

2159

2160

28. Change: The manufacturer of a molecular assay received clearance for a quantitative real-time PCR assay that included extraction kit reagents. The kit is therefore labeled for use with a set of extraction reagents. The manufacturer makes changes to the column substrate for the extraction method.

Relevant questions:

D1 – *Does the change alter the operating principle of the IVD?* No. The change in column substrate would not alter the operating principle.

D3 – *Does a risk assessment of the changed device identify any new risks or significantly modified existing risks?* Yes. The manufacturer’s risk assessment indicates that changing the column substrate could significantly change the analytical and clinical performance of the modified test compared to the previously cleared version of this device indicating new or significantly modified existing risks.

Decision: Submit the change in a new 510(k).

29. Change: The manufacturer of a bilirubin test system makes a change to the reagent, modifying from a liquid form to a lyophilized form of the reagent. The formulation and concentration of the reagent remain unchanged.

Relevant questions:

D1 – *Does the change alter the operating principle of the IVD?* No. This change in reagent would not alter the operating principle.

D3 – *Does a risk assessment of the changed device identify any new risks or significantly modified existing risks?* No. The manufacturer’s risk assessment indicates that the performance of the modified IVD could not significantly change from the previously cleared performance claims and that the modified IVD presents no new or significantly modified existing risks, since the change in reagent state does not change the concentration or formulation of the reagent.

D4 – *Do design verification and validation activities produce any unexpected issues of safety or effectiveness?* No. Standard methods and established and justified criteria are used to verify and validate the modification and results of the verification and validation studies do not indicate new issues of safety or effectiveness.

Decision: Document the change to file.

30. Change: The manufacturer makes a change in the traceability of an IVD calibrator.

Relevant questions:

Contains Nonbinding Recommendations

Draft – Not for Implementation

2161 D1- *Does the change alter the operating principle of the IVD?* No. A change in the
2162 traceability of an IVD calibrator would not alter the operating principle.
2163 D3 – *Does a risk assessment of the changed device identify any new risks or significantly*
2164 *modified existing risks?* Yes. The manufacturer’s risk assessment indicates that a change
2165 in the traceable reference standard for the assay calibrators could significantly change the
2166 clinical performance of the modified IVD from the previously cleared performance
2167 claims indicating new or significantly modified existing risks.
2168 **Decision:** Submit the change in a new 510(k).

2169
2170 **31. Change:** A manufacturer makes a change in the buffer solution of an IVD as a result of a
2171 change in vendor. The replacement buffer solution is equivalent to the previous buffer
2172 solution.

2173 **Relevant questions:**

2174 D1 – *Does the change alter the operating principle of the IVD?* No. The change in buffer
2175 solution would not alter the operating principle of the IVD.

2176 D3 – *Does a risk assessment of the changed device identify any new risks or significantly*
2177 *existing modified risks?* No. The manufacturer’s risk assessment indicates that the new
2178 buffer solution is equivalent to the previous buffer solution and indicates that the
2179 performance of the modified IVD could not significantly change from the previously
2180 cleared performance claims of the modified IVD or that the modified IVD presents new
2181 or significantly modified existing risks.

2182 D4 – *Do design verification and validation activities produce any unexpected issues of*
2183 *safety or effectiveness?* No. Standard methods and established and justified criteria are
2184 used to verify and validate the modification and results of the verification and validation
2185 studies do not indicate new issues of safety or effectiveness.

2186 **Decision:** Document the change to file.

2187
2188 **32. Change:** An IVD manufacturer makes a material change to their reagent and the
2189 manufacturer’s risk assessment indicates that the change in material could result in
2190 significantly changing the analytical performance from the previously cleared
2191 performance claims due to a potential change in the cut-off.

2192 **Relevant Questions:**

2193 D1 – *Does the change alter the operating principle of the IVD?* No. The change in
2194 material is not one that alters the operating principle of the IVD.

2195 D3 – *Does a risk assessment of the changed device identify any new risks or significantly*
2196 *modified existing risks?* Yes. The manufacturer’s risk assessment indicates that a change
2197 in the material of the reagent would result in a change in analytical cut-off that could
2198 significantly change the performance of the modified test compared to the previously
2199 cleared performance claims. In particular, this change in cut-off would be a change that
2200 is clinically significant in terms of clinical decision making since patients with samples
2201 around the cut-off could now receive a different diagnosis and treatment.

2202 **Decision:** Submit the change in a new 510(k).
2203

Contains Nonbinding Recommendations

Draft – Not for Implementation

2204 **33. Change:** A manufacturer changes the design of an IVD for diagnosing herpes simplex 1
2205 and 2 to a less strict performance specification that decreases both the sensitivity and
2206 specificity of the device to increase production.

2207 **Relevant questions:**

2208 D1 – *Does the change alter the operating principle of the IVD?* No. The change in
2209 design is not one that alters the operating principle of the IVD.

2210 D3 – *Does a risk assessment of the changed device identify any new risks or significantly*
2211 *modified existing risks?* Yes. The manufacturer’s risk assessment indicates that a change
2212 in the design of the IVD could significantly change the performance of the modified
2213 device compared to the previously cleared performance claims.

2214 **Decision:** Submit the change in a new 510(k).

2215

2216

2217

2218

DRAFT

Appendix B: Documentation

2219

2220

2221 Whenever a manufacturer changes its device, it must take certain actions to comply with the
2222 QS regulation, 21 CFR Part 820, unless a regulatory exemption exists. The QS regulation
2223 requires that design changes and production and process changes be documented prior to
2224 implementation. 21 CFR 820.30(i) and 820.70(b). If a manufacturer determines that the
2225 device modification(s) does not require a new 510(k), it should document the decision-
2226 making process and the basis for that conclusion. The documentation should be prepared in a
2227 way that an FDA investigator or other third party can understand what the change is and the
2228 rationale underlying the manufacturer's conclusion that a new 510(k) is not required.

2229

2230 FDA notes that only highlighting the flowcharts in this guidance document, or simply
2231 answering "yes" or "no" to each question without further details or justification, is not
2232 sufficient documentation. The manufacturer should provide robust justification of a decision
2233 that a new 510(k) is not required.

2234

2235 Documentation should include the following:

2236

2237

- Product name
- Date of modification assessment
- Description of the device
- Description of the modification(s)
- Reason why the modification(s) is being made
- Applicable regulatory history, including the 510(k) number of the last cleared version of the device
- Comparison of the modified device to the last cleared version of the device (consider including a table)
- Applicable elements of this guidance, including the applicable questions from the body of the document
- Analysis and assessment of the elements on this list and a conclusion of whether a new 510(k) is required
- Reference to related documents, particularly those that support the decision whether or not a new 510(k) is required (e.g., risk analysis)
- Signature(s)

2252

2253

2254

2255

2256

2257

2258

2259

2260

2261

It may be helpful to document the assessment of each modification in a way that corresponds to the decision-making framework discussed in this guidance document. If a manufacturer decides to do so, the documentation should list each relevant question, the answer to each of those questions, and the information and analysis that support the answer. The justification may be in the form of a detailed response, a relevant attachment, or other robust method that provides the rationale. Risk analyses will be particularly helpful in supporting the manufacturer's assessment. As a reminder, when making the decision on whether to submit a new 510(k), the manufacturer's basis for comparison of any changed device should be the

Contains Nonbinding Recommendations

Draft – Not for Implementation

2262 device described in the manufacturer's applicable most recently cleared 510(k), or to their
2263 legally marketed preamendments device.

2264

2265 Changes to a medical device or its processes vary in complexity. Some types of changes are
2266 straightforward and will generally result in a decision that a new 510(k) is not required. To
2267 that end, a manufacturer may establish a documentation process that accommodates different
2268 levels of documentation depending on the complexity of the change. Simple changes would
2269 have simple documentation and may not necessarily go through each question in detail; more
2270 complex changes should have more detailed documentation. Examples of types of changes
2271 that can typically be documented with simple documentation include:

2272

- 2273 • Modification of company labels to update to new company name, e.g., following
2274 acquisitions or address changes
- 2275 • Labeling layout changes where content is not changed, for instance, due to a
2276 corporate rebranding initiative
- 2277 • Addition of a unique device identifier (UDI) to labeling
- 2278 • Raw material supplier changes that only modify the reference number or brand name
2279 of raw materials and do not change the raw material itself

2280

2281 It is important that the manufacturer include, as part of the documentation process, a means
2282 to re-evaluate the change should initial assumptions subsequently not be met. In those
2283 situations, an update to the existing assessment, or a new assessment, should be documented.

2284

2285 The examples below are provided to illustrate one possible approach to documentation; other
2286 approaches may also be appropriate. Manufacturers are encouraged to use an approach that
2287 works for their specific purposes, taking into account the considerations discussed above.

2288 The first example below is a simple change that does not necessitate detailed analysis. The
2289 second example is a more complex change for which additional analysis and reference to
2290 supporting documentation are warranted. Note that these are generalized examples to
2291 demonstrate documentation principles and do not necessarily account for every possible
2292 detail, risk, or consideration.

2293

2294
2295
2296
2297
2298
2299
2300
2301
2302
2303
2304
2305
2306
2307
2308
2309
2310
2311
2312
2313
2314
2315
2316
2317
2318
2319
2320
2321
2322
2323
2324
2325
2326
2327
2328
2329
2330
2331
2332
2333
2334
2335
2336

Regulatory Change Assessment (Example 1)

Product Name: Device ABC

Date of Assessment: 10/25/16

Device Description: ABC is intended to treat headaches. Device consists of plates and screws. See design specifications at Document 15-XXXX.

Description of Change(s): ABC was recently acquired from Corporation X. Labeling will be updated to be consistent with our standard labeling. Specifically, the company logo, name, contact information, and labeling layout will be updated.

Reason for Change(s): To make ABC’s labeling consistent with our standard labeling.

Applicable Regulatory History (including 510(k) #s and comparison of modified device to last cleared version):

Device originally cleared in K10xxxx, cleared with updated plates in K12xxxx, cleared with updated screws in K14xxxx. Only changes between K14xxxx version and modified device are company logo, name, contact information, and labeling layout.

Completed Checklist Attached:

Yes

No (include rationale if selected)

The changes proposed are to the labeling, but do not change the content of the labeling aside from company name and contact information, which does not substantively affect the labeling and could not significantly affect safety or effectiveness. FDA’s Deciding When to Submit a 510(k) for a Change to an Existing Device guidance states at A4 that “Labeling changes that provide clarification without changing the meaning of the labeling would generally not result in the need to submit a new 510(k).”

Recommended Regulatory Action:

Submit 510(k)

Letter to file

Supporting Documents:

Design Specifications: 15-XXXX

Risk Assessment: N/A

Signatures: xxxx

Contains Nonbinding Recommendations

Draft – Not for Implementation

**Regulatory Change Assessment
(Example 2)**

2337
2338
2339
2340
2341
2342
2343
2344
2345
2346
2347
2348
2349
2350
2351
2352
2353
2354
2355
2356
2357
2358
2359
2360
2361
2362
2363
2364
2365
2366
2367
2368
2369
2370
2371
2372
2373
2374
2375
2376
2377
2378
2379
2380

Product Name: Cardiopulmonary Bypass (CPB) Cannula

Date of Assessment: 1/17/20

Device Description: Cardiopulmonary Bypass Cannula is intended to cannulate the vessels, perfuse the coronary arteries, and interconnect the catheters and cannulas with an oxygenator. The current design uses a 304 stainless steel guidewire with a coating composed of material X; the tips of the guidewire are partially uncoated. See design specifications at Document 18-XXXX.

Description of Change(s): The change is to remove the coating from the guidewire. Previously, the tips were uncoated, but now the entire guidewire will be uncoated. This modification applies to models 1 and 2. These models were originally cleared in K10xxxx. The uncoated guidewire will continue to be made of 304 stainless steel. The replacement and current guidewires are identical in design, performance, and materials, with the exception of the coating.

The current guidewire was chosen originally because it was from our current guidewire supplier (which supplies guidewires for other cannulas we manufacture), met the dimensional specifications, and was cost-effective. The coating on the original cannula was not a specific design feature that was required for the design, although it may contribute to longevity of the guidewire and enhances lubricity.

The proposed modification will remove the coating, which will expose the stainless steel along the entire length of the guidewire. This modification does not introduce any new blood-contacting materials as the current guidewire tip is uncoated, and was tested for biocompatibility in the original submission. We previously marketed a cannula with an uncoated 304 stainless steel guidewire, cleared in K08xxxx (see DHF XXXX).

Removing the coating from the guidewire will also result in a small change to the diameter of the guidewire due to the lack of the coating.

We have confirmed that the Type 304 material used for the uncoated guidewire is from the same supplier as we have used previously (see Communication 11/7/19-XXXX from supplier), and there have been no issues with rusting (which could introduce embolic particles during device use). In addition, we have confirmed that there are no manufacturing residuals on the surface of the Type 304 stainless steel guidewire that would be available to the patient now that the guidewire is no longer coated (see Memo 19-XXXX).

Reason for Change(s): The coated guidewire has been discontinued by the supplier.

Contains Nonbinding Recommendations

Draft – Not for Implementation

2381 **Applicable Regulatory History (including 510(k) #s and comparison of modified device**
2382 **to last cleared version):**
2383 CPB Cannula was originally cleared in K10xxxx. The labeling layout was changed in 2012
2384 (see Regulatory Change Assessment 12-XXXX). The differences between the K10xxxx
2385 version and the modified device therefore include an updated labeling layout and the removal
2386 of the guidewire coating.
2387

2388 **Completed Checklist Attached:**

- 2389 Yes
2390 No (include rationale if selected)

2391
2392 **Recommended Regulatory Action:**

- 2393 Submit 510(k)
2394 Letter to file
2395

2396 **Supporting Documents:**

- 2397 Design Specifications: 18-XXXX
2398 Risk Assessment: 20-XXXX
2399 Verification and Validation Summary: 20-YYYY
2400

2401 **Signatures:** xxxx
2402
2403

Contains Nonbinding Recommendations

Draft – Not for Implementation

2404 **Main Flowchart Questions**

2405 *Change made with intent to significantly improve the safety or effectiveness of the device,*
2406 *e.g., in response to a known risk, adverse event, etc.?*

2407 Yes

2408 No The change was made because the supplier discontinued the coating.

2409

2410 *Labeling change?*

2411 Yes

2412 No Labeling changes section N/A

2413

2414 *Technology or performance change?*

2415 Yes Coating will be removed which will change the design of the device and slightly
2416 decrease the diameter of the guidewire. This change will be evaluated to determine if
2417 this could affect the performance of the device.

2418 No

2419

2420 *Materials change?*

2421 Yes Removing the coating material from the device. This change will be evaluated to
2422 determine if processing could affect the biocompatibility of the device.

2423 No

2424

2425

Contains Nonbinding Recommendations

Draft – Not for Implementation

2426 **Labeling Questions**

2427

2428 *A1 – Is it a substantive change in the indications for use?*

2429 Yes Submit 510(k)

2430 No Go to A2

2431

2432 *A2 – Does the change add or delete a contraindication?*

2433 Yes Submit 510(k) (If adding a contraindication, submit CBE 510(k))

2434 No Go to A3

2435

2436 *A3 – Is it a change in warnings or precautions?*

2437 Yes Go to A5.1

2438 No Go to A4

2439

2440 *A4 – Does the change affect the instructions for use or other pieces of the labeling?*

2441 Yes Go to A5.1

2442 No Document to file

2443

2444 *A5.1 – Could the change affect the indications for use?*

2445 Yes Submit 510(k)

2446 No Go to A5.2

2447

2448 *A5.2 – Does a risk assessment of the changed device identify any new risks or significantly modified existing risks?*

2449 Yes Submit 510(k)

2450 No Document to file

2451

2452

2453

2454

Contains Nonbinding Recommendations

Draft – Not for Implementation

2455 **Technology, Engineering, and Performance Changes**

2456

2457 *B1 – Is the device an in vitro diagnostic device?*

2458 Yes Go to D1 (Technology, Engineering, Performance and Materials Changes for IVDs)

2459 No Go to B2

2460

2461 *B2 – Is it a control mechanism, operating principle, or energy type change?*

2462 Yes Submit 510(k)

2463 No Go to B3

2464

2465 *B3 – Is it a change in sterilization, cleaning, or disinfection?*

2466 Yes Go to B3.1

2467 No Go to B4

2468

2469 *B3.1 – Is it a change to an “established category B” or “novel” sterilization method, does the change lower the sterility assurance level, or is it a change to how the device is provided?*

2470 Yes Submit 510(k)

2471 No Go to B3.2

2472

2473 *B3.2 – Could the change significantly affect the performance or biocompatibility of the device?*

2474 Yes Submit 510(k)

2475 No Document to file

2476

2477 *B4 – Is there a change in packaging or expiration dating?*

2478 Yes Go to B4.1

2479 No Go to B5

2480

2481 *B4.1 – Is the same method or protocol, as described in a previously cleared 510(k), used to support the change?*

2482 Yes Document to file

2483 No Submit 510(k)

2484

2485 *B5 – Is it any other change in design (e.g., dimensions, performance specifications, wireless communication, components or accessories, or the patient/user interface)?*

2486 Yes Go to B5.1

2487 There are two changes, one to the coating of the guidewire, one to the dimensions of the guidewire. Each will be considered below.

2488

2489 No Document to file

2490

2491 *B5.1 – Does the change significantly affect the use of the device?*

2492

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 2498 Yes Submit 510(k)
2499 No Go to B5.2
2500 The lack of the coating and the small dimensional change are not expected to affect
2501 the use of the device.
2502
2503 *B5.2 – Does a risk assessment of the changed device identify any new risks or significantly*
2504 *modified existing risks?*
2505 Yes Submit 510(k)
2506 No Go to B5.3
2507 See full risk assessment in Document 20-XXXX.
2508 Dimensional change: it is unlikely that the small reduction in guidewire diameter
2509 could affect safety or effectiveness. Decreasing the diameter of the guidewire would
2510 not be expected to hinder the interaction between the guidewire, introducer, and
2511 cannula, and it would not be expected to reduce the strength of the guidewire, as the
2512 coating did not improve the strength of the wire and the wire itself remains
2513 unchanged.
2514
2515 Removal of the coating: it is unlikely, but possible, that the removal of the coating
2516 could impact the way the guidewire interacts with the introducer and cannula. We
2517 have previously obtained clearance for cannulas with uncoated stainless steel
2518 guidewires, however, which did not have markedly different performance (see DHF
2519 XXXX). This suggests that the significance of this change is low.
2520
2521 We have determined there are no new or significantly modified risks due to this
2522 change.
2523
2524 *B5.3 – Are clinical data necessary to evaluate safety or effectiveness for purposes of design*
2525 *validation?*
2526 Yes Submit 510(k)
2527 No Go to B5.4
2528
2529 *B5.4 – Do design verification and/or validation activities produce any unexpected issues of*
2530 *safety or effectiveness?*
2531 Yes Submit 510(k)
2532 No Document to file
2533 See verification and validation testing report in Document 20-YYYY, conducted after
2534 the risk assessment. Functional testing evaluated the interaction between the
2535 guidewire, introducer, and cannula to verify that the uncoated guidewire did not affect
2536 device performance. There were no unexpected issues of safety or effectiveness.
2537
2538
2539

Contains Nonbinding Recommendations

Draft – Not for Implementation

2540 **Materials Changes**

2541

2542 *C1 – Is the device an in vitro diagnostic product (IVD)?*

2543 Yes Go to D1 (Technology, Engineering, Performance and Materials Changes for IVDs)

2544 No Go to C2

2545

2546 *C2 – Is this a change in material type, material formulation, chemical composition, or the*
2547 *material's processing?*

2548 Yes Go to C3

2549 The coating material X will be removed.

2550

2551 No Document to file

2552

2553 *C3 – Will the changed material directly or indirectly contact body tissues or fluids?*

2554 Yes Go to C4

2555 No Go to C5

2556

2557 *C4 – Does a risk assessment identify any new or increased biocompatibility concerns?*

2558 Yes Go to C4.1

2559 No Go to C5

2560 The tips of the current guidewire are uncoated, so there is no new material here to
2561 create new biocompatibility concerns. The removal of the coating material is not
2562 expected to have a biocompatibility impact as the processing is unlikely to leave
2563 residuals that were previously masked by the coating. In addition, we have previously
2564 marketed cleared cannulas with uncoated stainless steel guidewires, which passed
2565 biocompatibility testing (see DHF XXXX). The source of the stainless steel used to
2566 manufacture these guidewires has not changed, and we have had no issues with
2567 rusting components, so embolic risk is not a concern.

2568

2569 *C4.1 – Has the manufacturer used the same material in a similar legally marketed device?*

2570 Yes Go to C5

2571 No Submit 510(k)

2572

2573 *C5 – Could the change affect the device's performance specifications?*

2574 Yes Go to B5

2575 See design change analysis above.

2576

2577 No Document to file

2578

2579

Contains Nonbinding Recommendations

Draft – Not for Implementation

2580 **Technology, Engineering, Performance, and Materials Changes for In Vitro Diagnostic**
2581 **Devices**

2582

2583 *D1 – Does the change alter the operating principle of the IVD?*

2584 Yes Submit 510(k)

2585 No Go to D2

2586

2587 *D2 – Is the change identified in a device-specific guidance or classification regulation?*

2588 Yes Submit 510(k)

2589 No Go to D3

2590

2591 *D3 – Does a risk assessment of the changed device identify any new risks or significantly*
2592 *modified existing risks?*

2593 Yes Submit 510(k)

2594 No Go to D4

2595

2596 *D4 – Do design verification and/or validation activities produce any unexpected issues of*
2597 *safety or effectiveness?*

2598 Yes Submit 510(k)

2599 No Document to file

2600

2601

DRAFT

Appendix C: Definitions

2602

2603

2604 The following definitions are provided to clarify the meaning of medical device terms used in
2605 this guidance document. Wherever possible, existing definitions from the FD&C Act,
2606 medical device regulations, or FDA guidance documents have been used. In some cases,
2607 where regulatory definitions are unavailable, we have relied on dictionary definitions of
2608 terms.

2609

2610 510(k) Holder: The person who possesses the 510(k) clearance for a device.

2611

2612 Contraindications: See “precautions, warnings and contraindications” below.

2613

2614 Control Mechanism: The manner by which the actions of a device are directed. An example
2615 of a change in control mechanism would be the replacement of an electromechanical control
2616 with a microprocessor control.

2617

2618 Dimensional Specifications: The physical size and shape of the device. Such specifications
2619 may include the length, width, thickness, or diameter of a device, as well as the location of a
2620 part or component of the device.

2621

2622 Documentation: For the purpose of this guidance, documentation means recording the
2623 rationale behind the manufacturer’s decision whether to submit a new 510(k) for changes in a
2624 device. Consideration of each decision point should be recorded, as well as the final
2625 conclusions reached. If testing or other engineering analysis is part of the process, the results
2626 of this activity should be recorded or referenced. A copy of this documentation should be
2627 maintained for future reference.

2628

2629 Energy Type, Character, or Source: The type of power input to or output from the device.
2630 Examples of a change in energy type or character would be a change from AC to battery
2631 power (input) or a change from ionizing radiation to ultrasound to measure a property of the
2632 body (output).

2633

2634 Environmental Specifications: The (range of) acceptable levels of environmental parameters
2635 or operating conditions under which the device will perform safely and effectively. Examples
2636 of changes in environmental specifications are expanding the acceptable temperature range in
2637 which the device will operate properly or hardening the device to significantly higher levels
2638 of electromagnetic interference.

2639

2640 Human Factors of Patient/User Interface: The human factors of the patient or user interface
2641 refer to the way in which the device and the patient or user interact. This includes the way in
2642 which the device presents alarms to the user, the layout of the control panel, the mode of
2643 presentation of information to the user or patient, and the way in which the device physically
2644 interacts with the user and/or patient (e.g., the way in which a CPAP mask attaches to a
2645 patient’s face, or the way a surgical instrument is designed to fit in a surgeon’s hand).

Contains Nonbinding Recommendations

Draft – Not for Implementation

2646

2647 Expiration Date: The date beyond which the product may cease to perform safely or
2648 effectively and beyond which the manufacturer states the product should not be used.

2649

2650 Harm: Physical injury or damage to the health of people.⁷

2651

2652 Hazard: Potential source of harm.

2653

2654 Intended Use: For purposes of substantial equivalence, the term “intended use” means the
2655 general purpose of the device or its function, and encompasses the indications for use.⁸

2656

2657 Indications for Use: The term indications for use, as defined in 21 CFR 814.20(b)(3)(i),
2658 describes the disease or condition the device will diagnose, treat, prevent, cure or mitigate,
2659 including a description of the patient population for which the device is intended.⁹

2660

2661 In Vitro Diagnostic Device: Those reagents, instruments, and systems intended for use in the
2662 diagnosis of disease or other conditions, including a determination of the state of health, in
2663 order to cure, mitigate, treat, or prevent disease or its sequelae. Such products are intended
2664 for use in the collection, preparation, and examination of specimens taken from the human
2665 body.¹⁰

2666

2667 Label: The term “label” means a display of written, printed, or graphic matter upon the
2668 immediate container of any article.¹¹

2669

2670 Labeling: The term “labeling” means all labels and other written, printed, or graphic matter
2671 (1) upon any article or its containers or wrappers, or (2) accompanying such article.¹² This
2672 can include, among other things, any user or maintenance manuals and, in some instances,
2673 promotional literature.

2674

2675 Manufacturer: For the purposes of this document, the term manufacturer includes any 510(k)
2676 holder, even if that person does not actually fabricate the existing device. The term also
2677 includes persons who have a preamendments device for a device type subject to premarket
2678 notification (510(k)).

2679

2680 Material Formulation: The base formulation of a polymer, alloy, etc., plus any additives,
2681 colors, etc., used to establish a property or the stability of the material. This does not include

⁷ Definition based on ISO 14971.

⁸ See FDA’s guidance [The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \(510\(k\)\)](#). See also 21 CFR 801.4.

⁹ See FDA’s guidance [The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \(510\(k\)\)](#).

¹⁰ 21 CFR 809.3(a).

¹¹ Section 201(k) of the FD&C Act.

¹² Section 201(m) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

2682 processing aids, mold release agents, residual contaminants, or other manufacturing aids that
2683 are not intended to be a part of the material, but that could be present as impurities on the
2684 final device. An example of a change in material formulation would be a change from a
2685 series 300 stainless steel to a series 400 stainless steel. Another example of a change in
2686 material formulation would be the addition or subtraction of a chemical or compound to or
2687 from a polymer.

2688
2689 Material Supplier: The firm supplying the raw material to a finished device manufacturer.

2690

2691 Material Type: The generic name of the material from which the device is manufactured. An
2692 example of a material type change would be the change from natural latex rubber to synthetic
2693 rubber.

2694

2695 Method of Sterilization: The physical or chemical mechanism used to achieve sterility or to
2696 achieve a specific sterility assurance level (SAL).

2697

2698 Operating Principle: The mode of operation or mechanism of action through which a device
2699 fulfills (or achieves) its intended use. An example of a change in operating principle would
2700 be using a new algorithm to compress images in a picture archiving and communications
2701 system. For an IVD, an example would be a change from immunofluorescence to ELISA.

2702

2703 Packaging: Any wrapping, containers, etc., used to protect, to preserve the sterility of, or to
2704 group medical devices.

2705

2706 Performance Specifications: The performance characteristics of a device as listed in device
2707 labeling or in finished product release specifications. Some examples of performance
2708 specifications are measurement accuracy, output accuracy, energy output level, and stability
2709 criteria.

2710

2711 Preamendments Device: A device commercially distributed in the United States prior to May
2712 28, 1976 that has not been significantly changed or modified since then, and for which
2713 premarket approval has not been required under section 515(b) of the FD&C Act.

2714

2715 Precautions, Warnings, and Contraindications:

2716

2717 • Precautions describe any special care to be exercised by a practitioner or patient for
2718 the safe and effective use of a device. This definition also includes limitations stated
2719 for IVDs.

2720 • Warnings describe serious adverse reactions and potential safety hazards that can
2721 occur in the proper use or misuse of a device, along with consequent limitations in
2722 use and mitigating steps to take if they occur.

2723

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 2724 • Contraindications describe situations in which the device should not be used because
2725 the risk of use clearly outweighs any reasonably foreseeable benefits.¹³
2726

2727 Reprocessing: Validated processes used to render a medical device, which has been
2728 previously used or contaminated, fit for a subsequent single use. These processes are
2729 designed to remove soil and contaminants by cleaning and to inactivate microorganisms by
2730 disinfection or sterilization.¹⁴

2731
2732 Reusable Medical Device: A device intended for repeated use either on the same or different
2733 patients, with appropriate cleaning and other reprocessing between uses.

2734
2735 Reuse: Use of a device more than once on a single patient or on more than one patient.
2736 Actions necessary for reuse of a device may include instructions for assembly/disassembly,
2737 on-site sterilization or disinfection, etc. This definition does not include the refurbishing or
2738 repair of a device for redistribution or resale.

2739
2740 Risk: The combination of the probability of occurrence of harm and the severity of that harm.
2741 For the purposes of this guidance, may relate to either safety or effectiveness (e.g., risk of
2742 decreasing device effectiveness).

2743
2744 Shelf-life: The term or period during which a device remains suitable for its intended use.
2745 This period ends at the device's expiration date.

2746
2747 Single-use Device (SUD): A device that is intended for one use or on a single patient during
2748 a single procedure.

2749
2750 Software: The set of electronic instructions used to control the actions or output of a medical
2751 device, to provide input to or output from a medical device, or to provide the actions of a
2752 medical device. This definition includes software that is embedded within or permanently a
2753 component of a medical device, software that is an accessory to another medical device, or
2754 software that is intended to be used for one or more medical purposes that performs these
2755 purposes without being part of a hardware medical device.

2756
2757 Sterility Assurance Level (SAL): The probability of a single viable microorganism occurring
2758 on an item after sterilization.

2759
2760 Sterilization: A validated process used to render product free from viable microorganisms.

2761 NOTE: In a sterilization process, the nature of microbial inactivation is described as
2762 exponential and, thus, the survival of a microorganism on an individual item can be

¹³ ODE Bluebook Memorandum G91-1, "[Device Labeling Guidance](#)."

¹⁴ See FDA's guidance [Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling](#).

Contains Nonbinding Recommendations

Draft – Not for Implementation

2763 expressed in terms of probability. While this probability can be reduced to a very low
2764 number, it can never be reduced to zero.¹⁵

2765

2766 User Interface: A device user interface includes all points of interaction between the product
2767 and the user, including elements such as displays, controls, packaging, product labels, and
2768 instructions for use.

2769

2770 Warnings: See “precautions, warnings, and contraindications” above.

2771

2772

2773

2774

2775

2776

2777

2778

2779

DRAFT

¹⁵ See FDA’s guidance [*Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling*](#).